

QC Chart of Internal Quality Control (IQC)for G6PD Quantitative Test

Select LotNo : AC1203N (2014-01-01 ~ 2100-12-31) [Change](#)

[Print Table](#)

Lab021

QC Control Lot No.	Control N		Control D	
	AC1203N		AC1203D	
Duration of the Analyzing	Month (2015/12)	CUM (2015/07/01~2015/12/31)	Month (2015/12)	CUM (2015/07/01~2015/12/31)
Runs (N)	4	26	4	26
Mean (U/gHb)	14.3	14.2	3.6	3.5
SD	0.2	0.2	0.2	0.1
CV (%)	1.4	1.4	5.6	2.9
Target Value (U/gHb)	14.2	14.2	3.3	3.3
Total Error (%)	3.5	2.8	20.2	11.8
TEa (%)	20	20	20	20
σ	>6	>6	1.9	4.8

Bias (%) = [(| Mean - Target |) / Target] x 100%
 TE : Total Error(%) = Bias (%) + 2 x CV (%)
 σ (Sigma) = [TEa% - Bias (%)] / CV (%)


Month : 2015 12 [Change](#) ; Cumulative : from 2015 07 01 to 2015 12 31 [Change](#)

[TOP](#)

Control N SDI QC Chart

Lot No.: AC1203N ; Duration : 2015-07-01 to 2015-12-31 ; Target : 14.2 ; SD : 2.6

Lab021




Month : 2015 12 [Change](#) ; Cumulative : from 2015 07 01 to 2015 12 31 [Change](#)

[TOP](#)

Control D SDI QC Chart

Lot No.: AC1203D ; Duration : 2015-07-01 to 2015-12-31 ; Target : 3.3 ; SD : 3.1

Lab021



Month : 2015 12 [Change](#) ; Cumulative : from 2015 07 01 to 2015 12 31 [Change](#)


[TOP](#)

Control N and Control D SDI QC Chart

Lot No.: AC1203N ; Duration : 2015-07-01 to 2015-12-31 ; Target : 14.2 ; SD : 2.6 (●)

Lot No.: AC1203D ; Duration : 2015-07-01 to 2015-12-31 ; Target : 3.3 ; SD : 3.1 (□)

Lab021



Month : 2015 12 ; Cumulative : from 2015 07 01 to 2015 12 31

[\[TOP\]](#)