

# Summary Report of IQC program for G6PD Quantitative Test - Medicon Group - September 2018 -


## I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AE0909N)	Control D (Lot No.:AE0909D)
Labs	8	8
Received results number (n)	50	50
Median	15.3 (U/gHb)	5.3 (U/gHb)
Mean	15.2 (U/gHb)	5.3 (U/gHb)
SD	1.2	0.4
CV	7.9%	7.5%
Range of G6PD	12.7 ~ 17.7 (U/gHb)	4.0 ~ 6.4 (U/gHb)
Range of Hb	1.5 ~ 2.3 (g/dL)	1.7 ~ 2.6 (g/dL)

\*The statistic results are calculated from all labs reported in this month

\*\* G6PD Method = Medicon reagent kit, 37°C

## II. The distribution of G6PD reported for each lab in this survey



# QC Chart of Internal Quality Control (IQC)for G6PD Quantitative Test

Select LotNo : AE0909N (2016-01-01 ~ 2100-12-31) [Change](#)

[Print Table](#)

## Lab040

QC Control Lot No.	Control N		Control D	
	AE0909N		AE0909D	
Duration of the Analyzing	Month (2018/09)	CUM (2018/04/17~2018/09/30)	Month (2018/09)	CUM (2018/04/17~2018/09/30)
Runs (N)	5	26	5	26
Mean (U/gHb)	13.9	13.6	4.6	4.7
SD	0.1	0.4	0.4	0.3
CV (%)	0.7	2.9	8.7	6.4
Target Value (U/gHb)	12.5	12.5	4.5	4.5
Total Error (%)	12.6	14.7	19.6	17.2
TEa (%)	20	20	20	20
$\sigma$	>6	3.9	2.0	2.4

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%  
 TE : Total Error(%) = Bias (%) + 2 x CV (%)  
 $\sigma$  (Sigma) = [TEa% - Bias (%) ] / CV (%)


Month : 2018 09 [Change](#) ; Cumulative : from 2018 04 17 to 2018 09 30 [Change](#)

[TOP](#)

## Control N SDI QC Chart

Lot No.: AE0909N ; Duration : 2018-04-17 to 2018-09-30 ; Target : 12.5 ; SD : 0.6

Lab040




Month : 2018 09 [Change](#) ; Cumulative : from 2018 04 17 to 2018 09 30 [Change](#)

[TOP](#)

## Control D SDI QC Chart

Lot No.: AE0909D ; Duration : 2018-04-17 to 2018-09-30 ; Target : 4.5 ; SD : 0.3

Lab040



Month : 2018 09 [Change](#) ; Cumulative : from 2018 04 17 to 2018 09 30 [Change](#)


[TOP](#)

## Control N and Control D SDI QC Chart

Lot No.: AE0909N ; Duration : 2018-04-17 to 2018-09-30 ; Target : 12.5 ; SD : 0.6 ( ● )

Lot No.: AE0909D ; Duration : 2018-04-17 to 2018-09-30 ; Target : 4.5 ; SD : 0.3 ( □ )

Lab040



Date 04-17 04-24 04-26 05-03 05-08 05-15 05-17 05-22 05-29 06-05 06-07 07-05 07-12 07-19 07-31 08-07 08-07 08-16 08-21 08-28 08-30 09-04 09-06 09-11 09-18 09-20

Month : 2018 09 Change ; Cumulative : from 2018 04 17 to 2018 09 30 Change

[TOP](#)

# Peer Group Statistics (Table 1)

Select LotNo :

Select Reagent Kit :

## Monthly

Month :

UnitID <sup>†</sup>	Reagent Kit (Code) <sup>†</sup>	Control N (Lot No.: AE0909N)								Control D (Lot No.: AE0909D)							
		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab018</a>	5	15.7	16.1	6	0.4	2.5	7.5	20	>6	5.6	5.6	6	0.1	1.8	3.6	20	>6
<a href="#">Lab019</a>	5	15.3	15.1	5	0.4	2.6	6.6	20	>6	5.3	5.3	5	0.1	1.9	3.8	20	>6
<a href="#">Lab026</a>	5	15.9	15.3	14	0.3	2.0	7.7	20	>6	5.8	5.3	14	0.4	7.5	23.7	20	1.5
<a href="#">Lab027</a>	5	11.5	17.1	8	0.5	2.9	54.5	20	-9.9	4.5	5.9	8	0.2	3.4	37.9	20	-3.3
<a href="#">Lab033</a>	5	14.4	15.1	4	0.4	2.6	10.2	20	5.8	5.8	5.3	4	0.1	1.9	12.4	20	6.0
<a href="#">Lab034</a>	5	14.4	14.2	4	0.5	3.5	8.4	20	5.3	5.8	5.1	4	0.1	2.0	16.0	20	4.0
<a href="#">Lab040</a>	5	12.5	13.9	5	0.1	0.7	12.6	20	>6	4.5	4.6	5	0.4	8.7	19.6	20	2.0
<a href="#">Lab043</a>	5	11.5	12.8	4	0.2	1.6	14.4	20	5.4	4.5	5.0	4	0.2	4.0	19.1	20	2.2
Total	-	-	15.2	50	1.2	7.9	-	-	-	-	5.3	50	0.4	7.5	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[\[TOP\]](#)

## Cumulative

Cumulative : from    to

UnitID <sup>†</sup>	Reagent Kit (Code) <sup>†</sup>	Control N (Lot No.: AE0909N)								Control D (Lot No.: AE0909D)							
		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab018</a>	5	15.7	16.1	60	0.7	4.3	11.2	20	4.1	5.6	5.7	60	0.3	5.3	12.3	20	3.4
<a href="#">Lab019</a>	5	15.3	15.1	34	0.4	2.6	6.6	20	>6	5.3	5.3	34	0.2	3.8	7.5	20	5.3
<a href="#">Lab026</a>	5	15.9	15.3	100	0.4	2.6	9.0	20	>6	5.8	5.2	100	0.5	9.6	29.6	20	1.0
<a href="#">Lab027</a>	5	11.5	16.1	52	1.4	8.7	57.4	20	-2.3	4.5	5.6	52	0.4	7.1	38.7	20	-0.6
<a href="#">Lab032</a>	5	15.9	15.8	22	1.0	6.3	13.3	20	3.1	5.8	5.7	22	0.4	7.0	15.8	20	2.6
<a href="#">Lab033</a>	5	14.4	14.8	23	0.7	4.7	12.2	20	3.7	5.8	5.2	23	0.2	3.8	18.0	20	2.5
<a href="#">Lab034</a>	5	14.4	14.1	28	0.5	3.5	9.2	20	5.1	5.8	5.0	28	0.3	6.0	25.8	20	1.0
<a href="#">Lab040</a>	5	12.5	13.2	38	0.6	4.5	14.7	20	3.2	4.5	4.7	38	0.3	6.4	17.2	20	2.4
<a href="#">Lab043</a>	5	11.5	13.4	33	1.2	9.0	34.4	20	0.4	4.5	5.1	33	0.3	5.9	25.1	20	1.1
Total	-	-	15.1	390	1.3	8.6	-	-	-	-	5.3	390	0.5	9.4	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[\[TOP\]](#)

Reagent Kit	Reagent Code
Medicon	5

## Peer Group Statistics (Table 2)

Select LotNo :

Select Reagent Kit :

### Control N Month vs. Cumulative

		Control N (Lot No.: AE0909N)															
		Month (2018/09)								CUM (2016/02/01~2018/09/30)							
UnitID <sup>↑</sup>	Reagent Kit (Code) <sup>↑</sup>	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab018</a>	5	15.7	16.1	6	0.4	2.5	7.5	20	>6	15.7	16.1	60	0.7	4.3	11.2	20	4.1
<a href="#">Lab019</a>	5	15.3	15.1	5	0.4	2.6	6.6	20	>6	15.3	15.1	34	0.4	2.6	6.6	20	>6
<a href="#">Lab026</a>	5	15.9	15.3	14	0.3	2.0	7.7	20	>6	15.9	15.3	100	0.4	2.6	9.0	20	>6
<a href="#">Lab027</a>	5	11.5	17.1	8	0.5	2.9	54.5	20	-9.9	11.5	16.1	52	1.4	8.7	57.4	20	-2.3
<a href="#">Lab032</a>	5	15.9	-	0	-	-	-	20	-	15.9	15.8	22	1.0	6.3	13.3	20	3.1
<a href="#">Lab033</a>	5	14.4	15.1	4	0.4	2.6	10.2	20	5.8	14.4	14.8	23	0.7	4.7	12.2	20	3.7
<a href="#">Lab034</a>	5	14.4	14.2	4	0.5	3.5	8.4	20	5.3	14.4	14.1	28	0.5	3.5	9.2	20	5.1
<a href="#">Lab040</a>	5	12.5	13.9	5	0.1	0.7	12.6	20	>6	12.5	13.2	38	0.6	4.5	14.7	20	3.2
<a href="#">Lab043</a>	5	11.5	12.8	4	0.2	1.6	14.4	20	5.4	11.5	13.4	33	1.2	9.0	34.4	20	0.4
Total	-	-	15.2	50	1.2	7.9	-	-	-	-	15.1	390	1.3	8.6	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month :

Cumulative : from    to

[TOP](#)

### Control D Month vs. Cumulative

		Control D (Lot No.: AE0909D)															
		Month (2018/09)								CUM (2016/02/01~2018/09/30)							
UnitID <sup>↑</sup>	Reagent Kit (Code) <sup>↑</sup>	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab018</a>	5	5.6	5.6	6	0.1	1.8	3.6	20	>6	5.6	5.7	60	0.3	5.3	12.3	20	3.4
<a href="#">Lab019</a>	5	5.3	5.3	5	0.1	1.9	3.8	20	>6	5.3	5.3	34	0.2	3.8	7.5	20	5.3
<a href="#">Lab026</a>	5	5.8	5.3	14	0.4	7.5	23.7	20	1.5	5.8	5.2	100	0.5	9.6	29.6	20	1.0
<a href="#">Lab027</a>	5	4.5	5.9	8	0.2	3.4	37.9	20	-3.3	4.5	5.6	52	0.4	7.1	38.7	20	-0.6
<a href="#">Lab032</a>	5	5.8	-	0	-	-	-	20	-	5.8	5.7	22	0.4	7.0	15.8	20	2.6
<a href="#">Lab033</a>	5	5.8	5.3	4	0.1	1.9	12.4	20	6.0	5.8	5.2	23	0.2	3.8	18.0	20	2.5
<a href="#">Lab034</a>	5	5.8	5.1	4	0.1	2.0	16.0	20	4.0	5.8	5.0	28	0.3	6.0	25.8	20	1.0
<a href="#">Lab040</a>	5	4.5	4.6	5	0.4	8.7	19.6	20	2.0	4.5	4.7	38	0.3	6.4	17.2	20	2.4
<a href="#">Lab043</a>	5	4.5	5.0	4	0.2	4.0	19.1	20	2.2	4.5	5.1	33	0.3	5.9	25.1	20	1.1
Total	-	-	5.3	50	0.4	7.5	-	-	-	-	5.3	390	0.5	9.4	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month :

Cumulative : from    to

[TOP](#)

Reagent Kit	Reagent Code
Medicon	5