

# Summary Report of IQC program for G6PD Quantitative Test - Medicon Group - April 2023 -


## I. The statistic results of all laboratories in this month

| G6PD                        | Control N<br>(Lot No.:BJ0922N) | Control D<br>(Lot No.: BJ0922D) |
|-----------------------------|--------------------------------|---------------------------------|
| Labs                        | 9                              | 9                               |
| Received results number (n) | 30                             | 30                              |
| Median                      | 14.3 (U/g Hb)                  | 4.7 (U/g Hb)                    |
| Mean                        | 14.1 (U/g Hb)                  | 4.7 (U/g Hb)                    |
| SD                          | 0.5                            | 0.3                             |
| CV                          | 3.5%                           | 4.3%                            |
| Range of G6PD               | 12.8 ~ 14.9 (U/g Hb)           | 4.3 ~ 5.3 (U/g Hb)              |
| Range of Hb                 | 2.1 ~ 2.8 (g/dL)               | 1.9 ~ 2.4 (g/dL)                |

\*The statistic results are calculated from all labs reported in this month

\*\* G6PD Method = Medicon reagent kit, 37°C

## II. The distribution of G6PD reported for each lab in this survey



# QC Chart of Internal Quality Control (IQC) for G6PD Quantitative Test

Select LotNo : BJ0922N (2021-01-01 ~ 2100-12-31) Change

[Print Table](#)

## Lab040

| QC Control Lot No.        | Control N       |                             | Control D       |                             |
|---------------------------|-----------------|-----------------------------|-----------------|-----------------------------|
|                           | BJ0922N         |                             | BJ0922D         |                             |
| Duration of the Analyzing | Month (2023/04) | CUM (2022/11/08~2023/04/30) | Month (2023/04) | CUM (2022/11/08~2023/04/30) |
| Runs (N)                  | 3               | 26                          | 3               | 26                          |
| Mean (U/gHb)              | 14.7            | 14.8                        | 4.7             | 4.7                         |
| SD                        | 0.2             | 0.2                         | 0.1             | 0.1                         |
| CV (%)                    | 1.4             | 1.4                         | 2.1             | 2.1                         |
| Target Value (U/gHb)      | 14.4            | 14.4                        | 4.6             | 4.6                         |
| Total Error (%)           | 4.8             | 5.5                         | 6.4             | 6.4                         |
| TEa (%)                   | 20              | 20                          | 20              | 20                          |
| $\sigma$                  | >6              | >6                          | >6              | >6                          |

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%  
 TE : Total Error(%) = Bias (%) + 2 x CV (%)  
 $\sigma$  (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month : 2023 04 Change ; Cumulative : from 2022 11 08 to 2023 04 30 Change


[\[TOP\]](#)

## Control N and Control D SDI QC Chart

Lot No.: BJ0922N ; Duration : 2022-11-08 to 2023-04-30 ; Target : 14.4 ; SD : 0.80 (●)

Lot No.: BJ0922D ; Duration : 2022-11-08 to 2023-04-30 ; Target : 4.6 ; SD : 0.20 (□)

Lab040



Date 11-08 11-11 11-18 11-25 12-02 12-09 12-14 12-16 12-23 12-29 01-06 01-13 01-20 01-27 02-03 02-10 02-17 02-24 03-03 03-10 03-17 03-24 03-30 04-06 04-14 04-21


Month : 2023 04 Change ; Cumulative : from 2022 11 08 to 2023 04 30 Change

[\[TOP\]](#)

## Control N SDI QC Chart

Lot No.: BJ0922N ; Duration : 2022-11-08 to 2023-04-30 ; Target : 14.4 ; SD : 0.80

Lab040



Date 11-08 11-11 11-18 11-25 12-02 12-09 12-14 12-16 12-23 12-29 01-06 01-13 01-20 01-27 02-03 02-10 02-17 02-24 03-03 03-10 03-17 03-24 03-30 04-06 04-14 04-21


Month : 2023 04 Change ; Cumulative : from 2022 11 08 to 2023 04 30 Change

[\[TOP\]](#)

## Control D SDI QC Chart

Lot No.: BJ0922D ; Duration : 2022-11-08 to 2023-04-30 ; Target : 4.6 ; SD : 0.20

Lab040



**Date** 11-08 11-11 11-18 11-25 12-02 12-09 12-14 12-16 12-23 12-29 01-06 01-13 01-20 01-27 02-03 02-10 02-17 02-24 03-03 03-10 03-17 03-24 03-30 04-06 04-14 04-21

Month :    ; Cumulative : from    to

[TOP](#)

# Peer Group Statistics (Table 1)

Select LotNo : BJ0922N (2021-01-01 ~ 2100-12-31) Change

Select Reagent Kit : 5 - Medicon Change

[Print Table 1](#)

## Monthly

Month : 2023 04 Change

| UnitID <span>↑</span>  | Reagent Kit (Code) <span>↑</span> | Control N (Lot No.: BJ0922N) |              |            |     |        |        |         |     | Control D (Lot No.: BJ0922D) |              |            |     |        |        |         |     |
|------------------------|-----------------------------------|------------------------------|--------------|------------|-----|--------|--------|---------|-----|------------------------------|--------------|------------|-----|--------|--------|---------|-----|
|                        |                                   | Target (U/gHb)               | Mean (U/gHb) | n for Mean | SD  | CV (%) | TE (%) | TEa (%) | σ   | Target (U/gHb)               | Mean (U/gHb) | n for Mean | SD  | CV (%) | TE (%) | TEa (%) | σ   |
| <a href="#">Lab019</a> | 5                                 | 14.3                         | 13.8         | 4          | 0.8 | 5.8    | 15.1   | 20      | 2.8 | 4.6                          | 4.6          | 4          | 0.1 | 2.2    | 4.3    | 20      | >6  |
| <a href="#">Lab020</a> | 5                                 | 14.4                         | 14.6         | 1          | -   | -      | -      | 20      | -   | 4.6                          | 4.8          | 1          | -   | -      | -      | 20      | -   |
| <a href="#">Lab022</a> | 5                                 | 14.4                         | 14.4         | 5          | 0.1 | 0.7    | 1.4    | 20      | >6  | 4.6                          | 4.9          | 5          | 0.4 | 8.2    | 22.8   | 20      | 1.6 |
| <a href="#">Lab032</a> | 5                                 | 14.9                         | 14.0         | 4          | 0.5 | 3.6    | 13.2   | 20      | 3.9 | 4.8                          | 4.4          | 4          | 0.1 | 2.3    | 12.9   | 20      | 5.1 |
| <a href="#">Lab033</a> | 5                                 | 13.8                         | 13.4         | 4          | 0.2 | 1.5    | 5.9    | 20      | >6  | 4.7                          | 4.7          | 4          | 0.1 | 2.1    | 4.3    | 20      | >6  |
| <a href="#">Lab037</a> | 5                                 | 13.8                         | 14.3         | 1          | -   | -      | -      | 20      | -   | 4.7                          | 4.4          | 1          | -   | -      | -      | 20      | -   |
| <a href="#">Lab040</a> | 5                                 | 14.4                         | 14.7         | 3          | 0.2 | 1.4    | 4.8    | 20      | >6  | 4.6                          | 4.7          | 3          | 0.1 | 2.1    | 6.4    | 20      | >6  |
| <a href="#">Lab043</a> | 5                                 | 13.2                         | 13.8         | 4          | 0.3 | 2.2    | 8.9    | 20      | >6  | 4.2                          | 4.7          | 4          | 0.1 | 2.1    | 16.2   | 20      | 3.9 |
| <a href="#">Lab049</a> | 5                                 | 13.8                         | 14.5         | 4          | 0.3 | 2.1    | 9.2    | 20      | >6  | 4.7                          | 4.9          | 4          | 0.3 | 6.1    | 16.5   | 20      | 2.6 |
| Total                  | -                                 | -                            | 14.1         | 30         | 0.5 | 3.5    | -      | -       | -   | -                            | 4.7          | 30         | 0.3 | 6.4    | -      | -       | -   |

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[\[TOP\]](#)

## Cumulative

Cumulative : from 2021 01 01 to 2023 04 30 Change

| UnitID <span>↑</span>  | Reagent Kit (Code) <span>↑</span> | Control N (Lot No.: BJ0922N) |              |            |     |        |        |         |     | Control D (Lot No.: BJ0922D) |              |            |     |        |        |         |     |
|------------------------|-----------------------------------|------------------------------|--------------|------------|-----|--------|--------|---------|-----|------------------------------|--------------|------------|-----|--------|--------|---------|-----|
|                        |                                   | Target (U/gHb)               | Mean (U/gHb) | n for Mean | SD  | CV (%) | TE (%) | TEa (%) | σ   | Target (U/gHb)               | Mean (U/gHb) | n for Mean | SD  | CV (%) | TE (%) | TEa (%) | σ   |
| <a href="#">Lab018</a> | 5                                 | 16.2                         | 16.4         | 8          | 0.4 | 2.4    | 6.1    | 20      | >6  | 5.5                          | 5.5          | 8          | 0.2 | 3.6    | 7.3    | 20      | 5.6 |
| <a href="#">Lab019</a> | 5                                 | 14.3                         | 14.3         | 54         | 0.6 | 4.2    | 8.4    | 20      | 4.8 | 4.6                          | 4.6          | 54         | 0.2 | 4.3    | 8.7    | 20      | 4.7 |
| <a href="#">Lab020</a> | 5                                 | 14.4                         | 14.4         | 60         | 0.5 | 3.5    | 6.9    | 20      | 5.7 | 4.6                          | 4.7          | 60         | 0.1 | 2.1    | 6.4    | 20      | >6  |
| <a href="#">Lab022</a> | 5                                 | 14.4                         | 14.4         | 27         | 0.5 | 3.5    | 6.9    | 20      | 5.7 | 4.6                          | 4.7          | 27         | 0.3 | 6.4    | 14.9   | 20      | 2.8 |
| <a href="#">Lab028</a> | 5                                 | 15.2                         | 15.1         | 34         | 1.1 | 7.3    | 15.2   | 20      | 2.6 | 4.9                          | 4.9          | 34         | 0.3 | 6.1    | 12.2   | 20      | 3.3 |
| <a href="#">Lab032</a> | 5                                 | 14.9                         | 14.3         | 123        | 0.4 | 2.8    | 9.6    | 20      | 5.7 | 4.8                          | 4.5          | 123        | 0.2 | 4.4    | 15.1   | 20      | 3.1 |
| <a href="#">Lab033</a> | 5                                 | 13.8                         | 13.7         | 68         | 0.6 | 4.4    | 9.5    | 20      | 4.4 | 4.7                          | 4.6          | 68         | 0.3 | 6.5    | 15.2   | 20      | 2.7 |
| <a href="#">Lab037</a> | 5                                 | 13.8                         | 14.5         | 53         | 0.5 | 3.4    | 12.0   | 20      | 4.4 | 4.7                          | 4.7          | 53         | 0.2 | 4.3    | 8.5    | 20      | 4.7 |
| <a href="#">Lab040</a> | 5                                 | 14.4                         | 14.7         | 47         | 0.2 | 1.4    | 4.8    | 20      | >6  | 4.6                          | 4.6          | 47         | 0.1 | 2.2    | 4.3    | 20      | >6  |
| <a href="#">Lab043</a> | 5                                 | 13.2                         | 13.7         | 85         | 1.3 | 9.5    | 22.8   | 20      | 1.7 | 4.2                          | 4.5          | 85         | 0.5 | 11.1   | 29.4   | 20      | 1.2 |
| <a href="#">Lab049</a> | 5                                 | 13.8                         | 14.5         | 91         | 0.7 | 4.8    | 14.7   | 20      | 3.1 | 4.7                          | 4.7          | 91         | 0.2 | 4.3    | 8.5    | 20      | 4.7 |
| Total                  | -                                 | -                            | 14.3         | 650        | 0.8 | 5.6    | -      | -       | -   | -                            | 4.6          | 650        | 0.3 | 6.5    | -      | -       | -   |

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[\[TOP\]](#)

| Reagent Kit | Reagent Code |
|-------------|--------------|
| Medicon     | 5            |

## Peer Group Statistics (Table 2)

Select LotNo : BJ0922N (2021-01-01 ~ 2100-12-31) Change

Select Reagent Kit : 5 - Medicon Change

[Print Table 2](#)

### Control N Month vs. Cumulative

|                        |                      | Control N (Lot No.: BJ0922N) |              |            |     |        |        |         |     |                             |              |            |     |        |        |         |     |
|------------------------|----------------------|------------------------------|--------------|------------|-----|--------|--------|---------|-----|-----------------------------|--------------|------------|-----|--------|--------|---------|-----|
|                        |                      | Month (2023/04)              |              |            |     |        |        |         |     | CUM (2021/01/01~2023/04/30) |              |            |     |        |        |         |     |
| UnitID ↑               | Reagent Kit (Code) ↑ | Target (U/gHb)               | Mean (U/gHb) | n for Mean | SD  | CV (%) | TE (%) | TEa (%) | σ   | Target (U/gHb)              | Mean (U/gHb) | n for Mean | SD  | CV (%) | TE (%) | TEa (%) | σ   |
| <a href="#">Lab018</a> | 5                    | 16.2                         | -            | 0          | -   | -      | -      | 20      | -   | 16.2                        | 16.4         | 8          | 0.4 | 2.4    | 6.1    | 20      | >6  |
| <a href="#">Lab019</a> | 5                    | 14.3                         | 13.8         | 4          | 0.8 | 5.8    | 15.1   | 20      | 2.8 | 14.3                        | 14.3         | 54         | 0.6 | 4.2    | 8.4    | 20      | 4.8 |
| <a href="#">Lab020</a> | 5                    | 14.4                         | 14.6         | 1          | -   | -      | -      | 20      | -   | 14.4                        | 14.4         | 60         | 0.5 | 3.5    | 6.9    | 20      | 5.7 |
| <a href="#">Lab022</a> | 5                    | 14.4                         | 14.4         | 5          | 0.1 | 0.7    | 1.4    | 20      | >6  | 14.4                        | 14.4         | 27         | 0.5 | 3.5    | 6.9    | 20      | 5.7 |
| <a href="#">Lab028</a> | 5                    | 15.2                         | -            | 0          | -   | -      | -      | 20      | -   | 15.2                        | 15.1         | 34         | 1.1 | 7.3    | 15.2   | 20      | 2.6 |
| <a href="#">Lab032</a> | 5                    | 14.9                         | 14.0         | 4          | 0.5 | 3.6    | 13.2   | 20      | 3.9 | 14.9                        | 14.3         | 123        | 0.4 | 2.8    | 9.6    | 20      | 5.7 |
| <a href="#">Lab033</a> | 5                    | 13.8                         | 13.4         | 4          | 0.2 | 1.5    | 5.9    | 20      | >6  | 13.8                        | 13.7         | 68         | 0.6 | 4.4    | 9.5    | 20      | 4.4 |
| <a href="#">Lab037</a> | 5                    | 13.8                         | 14.3         | 1          | -   | -      | -      | 20      | -   | 13.8                        | 14.5         | 53         | 0.5 | 3.4    | 12.0   | 20      | 4.4 |
| <a href="#">Lab040</a> | 5                    | 14.4                         | 14.7         | 3          | 0.2 | 1.4    | 4.8    | 20      | >6  | 14.4                        | 14.7         | 47         | 0.2 | 1.4    | 4.8    | 20      | >6  |
| <a href="#">Lab043</a> | 5                    | 13.2                         | 13.8         | 4          | 0.3 | 2.2    | 8.9    | 20      | >6  | 13.2                        | 13.7         | 85         | 1.3 | 9.5    | 22.8   | 20      | 1.7 |
| <a href="#">Lab049</a> | 5                    | 13.8                         | 14.5         | 4          | 0.3 | 2.1    | 9.2    | 20      | >6  | 13.8                        | 14.5         | 91         | 0.7 | 4.8    | 14.7   | 20      | 3.1 |
| Total                  | -                    | -                            | 14.1         | 30         | 0.5 | 3.5    | -      | -       | -   | -                           | 14.3         | 650        | 0.8 | 5.6    | -      | -       | -   |

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month : 2023 04 Change

Cumulative : from 2021 01 01 to 2023 04 30 Change

[TOP](#)

### Control D Month vs. Cumulative

|                        |                      | Control D (Lot No.: BJ0922D) |              |            |     |        |        |         |     |                             |              |            |     |        |        |         |     |
|------------------------|----------------------|------------------------------|--------------|------------|-----|--------|--------|---------|-----|-----------------------------|--------------|------------|-----|--------|--------|---------|-----|
|                        |                      | Month (2023/04)              |              |            |     |        |        |         |     | CUM (2021/01/01~2023/04/30) |              |            |     |        |        |         |     |
| UnitID ↑               | Reagent Kit (Code) ↑ | Target (U/gHb)               | Mean (U/gHb) | n for Mean | SD  | CV (%) | TE (%) | TEa (%) | σ   | Target (U/gHb)              | Mean (U/gHb) | n for Mean | SD  | CV (%) | TE (%) | TEa (%) | σ   |
| <a href="#">Lab018</a> | 5                    | 5.5                          | -            | 0          | -   | -      | -      | 20      | -   | 5.5                         | 5.5          | 8          | 0.2 | 3.6    | 7.3    | 20      | 5.6 |
| <a href="#">Lab019</a> | 5                    | 4.6                          | 4.6          | 4          | 0.1 | 2.2    | 4.3    | 20      | >6  | 4.6                         | 4.6          | 54         | 0.2 | 4.3    | 8.7    | 20      | 4.7 |
| <a href="#">Lab020</a> | 5                    | 4.6                          | 4.8          | 1          | -   | -      | -      | 20      | -   | 4.6                         | 4.7          | 60         | 0.1 | 2.1    | 6.4    | 20      | >6  |
| <a href="#">Lab022</a> | 5                    | 4.6                          | 4.9          | 5          | 0.4 | 8.2    | 22.8   | 20      | 1.6 | 4.6                         | 4.7          | 27         | 0.3 | 6.4    | 14.9   | 20      | 2.8 |
| <a href="#">Lab028</a> | 5                    | 4.9                          | -            | 0          | -   | -      | -      | 20      | -   | 4.9                         | 4.9          | 34         | 0.3 | 6.1    | 12.2   | 20      | 3.3 |
| <a href="#">Lab032</a> | 5                    | 4.8                          | 4.4          | 4          | 0.1 | 2.3    | 12.9   | 20      | 5.1 | 4.8                         | 4.5          | 123        | 0.2 | 4.4    | 15.1   | 20      | 3.1 |
| <a href="#">Lab033</a> | 5                    | 4.7                          | 4.7          | 4          | 0.1 | 2.1    | 4.3    | 20      | >6  | 4.7                         | 4.6          | 68         | 0.3 | 6.5    | 15.2   | 20      | 2.7 |
| <a href="#">Lab037</a> | 5                    | 4.7                          | 4.4          | 1          | -   | -      | -      | 20      | -   | 4.7                         | 4.7          | 53         | 0.2 | 4.3    | 8.5    | 20      | 4.7 |
| <a href="#">Lab040</a> | 5                    | 4.6                          | 4.7          | 3          | 0.1 | 2.1    | 6.4    | 20      | >6  | 4.6                         | 4.6          | 47         | 0.1 | 2.2    | 4.3    | 20      | >6  |
| <a href="#">Lab043</a> | 5                    | 4.2                          | 4.7          | 4          | 0.1 | 2.1    | 16.2   | 20      | 3.9 | 4.2                         | 4.5          | 85         | 0.5 | 11.1   | 29.4   | 20      | 1.2 |
| <a href="#">Lab049</a> | 5                    | 4.7                          | 4.9          | 4          | 0.3 | 6.1    | 16.5   | 20      | 2.6 | 4.7                         | 4.7          | 91         | 0.2 | 4.3    | 8.5    | 20      | 4.7 |
| Total                  | -                    | -                            | 4.7          | 30         | 0.3 | 6.4    | -      | -       | -   | -                           | 4.6          | 650        | 0.3 | 6.5    | -      | -       | -   |

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month : 2023 04 Change

Cumulative : from 2021 01 01 to 2023 04 30 Change

[TOP](#)

| Reagent Kit | Reagent Code |
|-------------|--------------|
| Medicon     | 5            |