

Summary Report of IQC program for G6PD Quantitative Test - AMP Group

- June 2017 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	8	8
Received results number (n)	43	43
Median	11.5 (U/gHb)	2.7 (U/gHb)
Mean	11.5 (U/gHb)	2.7 (U/gHb)
SD	0.8	0.3
CV	7.0%	11.1%
Range of G6PD	9.7 ~ 13.3 (U/gHb)	2.1 ~ 3.2 (U/gHb)
Range of Hb	1.8 ~ 2.8 (g/dL)	1.6 ~ 3.3 (g/dL)

*The statistic results are calculated from all labs reported in this month

** G6PD Method = AMP reagent kit, 37°C

II. The distribution of G6PD reported for each lab in this survey

