

Summary Report of IQC program for G6PD Quantitative Test - Medicon Group

- April 2024 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:BJ0922N)	Control D (Lot No.: BJ0922D)
	Labs	7
Received results number (n)	37	37
Median	14.6 (U/g Hb)	4.9 (U/g Hb)
Mean	14.6 (U/g Hb)	5.0 (U/g Hb)
SD	0.5	0.3
CV	3.4%	6.0%
Range of G6PD	13.1 ~ 15.4 (U/g Hb)	4.6 ~ 5.7 (U/g Hb)
Range of Hb	2.0 ~ 2.7 (g/dL)	1.8 ~ 2.3 (g/dL)

*The statistic results are calculated from all labs reported in this month

** G6PD Method = Medicon reagent kit, 37°C

II. The distribution of G6PD reported for each lab in this survey

