

# QC Chart of Internal Quality Control (IQC)for G6PD Quantitative Test

Select LotNo : AC1203N (2014-01-01 ~ 2100-12-31) [Change](#)

[Print Table](#)

## Lab028

QC Control Lot No.	Control N		Control D	
	AC1203N		AC1203D	
Duration of the Analyzing	Month (2018/08)	CUM (2018/07/01~2018/08/31)	Month (2018/08)	CUM (2018/07/01~2018/08/31)
Runs (N)	13	26	13	26
Mean (U/gHb)	15.0	15.1	3.6	3.6
SD	0.9	0.9	0.2	0.2
CV (%)	6.0	6.0	5.6	5.6
Target Value (U/gHb)	14.2	14.2	3.3	3.3
Total Error (%)	17.6	18.3	20.2	20.2
TEa (%)	20	20	20	20
$\sigma$	2.4	2.3	1.9	1.9

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%  
 TE : Total Error(%) = Bias (%) + 2 x CV (%)  
 $\sigma$  (Sigma) = [TEa% - Bias (%) ] / CV (%)


Month : 2018 08 [Change](#) ; Cumulative : from 2018 07 01 to 2018 08 31 [Change](#)

[TOP](#)

## Control N SDI QC Chart

Lot No.: AC1203N ; Duration : 2018-07-01 to 2018-08-31 ; Target : 14.2 ; SD : 2.6

Lab028




Month : 2018 08 [Change](#) ; Cumulative : from 2018 07 01 to 2018 08 31 [Change](#)

[TOP](#)

## Control D SDI QC Chart

Lot No.: AC1203D ; Duration : 2018-07-01 to 2018-08-31 ; Target : 3.3 ; SD : 3.1

Lab028



Month : 2018 08 [Change](#) ; Cumulative : from 2018 07 01 to 2018 08 31 [Change](#)


[TOP](#)

# Control N and Control D SDI QC Chart

Lot No.: AC1203N ; Duration : 2018-07-01 to 2018-08-31 ; Target : 14.2 ; SD : 2.6 ( ● )

Lot No.: AC1203D ; Duration : 2018-07-01 to 2018-08-31 ; Target : 3.3 ; SD : 3.1 ( □ )

Lab028



Month : 2018 08  ; Cumulative : from 2018 07 01 to 2018 08 31

[TOP](#)

# Peer Group Statistics (Table 1)

Select LotNo :

Select Reagent Kit :

## Monthly

Month :

UnitID ↑	Reagent Kit (Code) ↑	Control N (Lot No.: AC1203N)								Control D (Lot No.: AC1203D)							
		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab022</a>	5	14.4	14.1	5	0.4	2.8	7.8	20	>6	3.3	3.2	5	0.3	9.4	21.8	20	1.8
<a href="#">Lab028</a>	5	14.2	15.0	13	0.9	6.0	17.6	20	2.4	3.3	3.6	13	0.2	5.6	20.2	20	1.9
Total	-	-	14.8	18	0.9	6.1	-	-	-	-	3.5	18	0.3	8.6	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[\[TOP\]](#)

## Cumulative

Cumulative : from    to

UnitID ↑	Reagent Kit (Code) ↑	Control N (Lot No.: AC1203N)								Control D (Lot No.: AC1203D)							
		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab019</a>	5	14.8	14.9	30	0.7	4.7	10.1	20	4.1	3.4	3.6	30	0.6	16.7	39.2	20	0.8
<a href="#">Lab020</a>	5	14.2	15.7	29	0.8	5.1	20.8	20	1.9	3.3	3.5	29	0.2	5.7	17.5	20	2.4
<a href="#">Lab021</a>	5	12.3	15.5	12	0.5	3.2	32.5	20	-1.9	2.8	4.5	12	1.1	24.4	109.6	20	-1.7
<a href="#">Lab022</a>	5	14.4	14.6	18	0.5	3.4	8.2	20	5.5	3.3	3.3	18	0.2	6.1	12.1	20	3.3
<a href="#">Lab026</a>	5	14.6	14.7	16	0.5	3.4	7.5	36	>6	3.2	3.2	16	0.1	3.1	6.3	36	>6
<a href="#">Lab028</a>	5	14.2	15.1	102	0.8	5.3	16.9	20	2.6	3.3	3.6	102	0.2	5.6	20.2	20	1.9
<a href="#">Lab037</a>	5	12.6	15.7	15	0.5	3.2	31.0	20	-1.4	3.0	3.5	15	0.1	2.9	22.4	20	1.1
<a href="#">Lab041</a>	5	12.8	12.9	1	-	-	-	20	-	2.9	3.2	1	-	-	-	20	-
Total	-	-	15.1	223	0.8	5.3	-	-	-	-	3.6	223	0.4	11.1	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[\[TOP\]](#)

Reagent Kit	Reagent Code
Medicon	5

## Peer Group Statistics (Table 2)

Select LotNo :

Select Reagent Kit :

### Control N Month vs. Cumulative

		Control N (Lot No.: AC1203N)															
		Month (2018/08)								CUM (2014/02/01~2018/08/31)							
UnitID <span style="font-size: small;">↑</span>	Reagent Kit (Code) <span style="font-size: small;">↑</span>	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab019</a>	5	14.8	-	0	-	-	-	20	-	14.8	14.9	30	0.7	4.7	10.1	20	4.1
<a href="#">Lab020</a>	5	14.2	-	0	-	-	-	20	-	14.2	15.7	29	0.8	5.1	20.8	20	1.9
<a href="#">Lab021</a>	5	12.3	-	0	-	-	-	20	-	12.3	15.5	12	0.5	3.2	32.5	20	-1.9
<a href="#">Lab022</a>	5	14.4	14.1	5	0.4	2.8	7.8	20	>6	14.4	14.6	18	0.5	3.4	8.2	20	5.5
<a href="#">Lab026</a>	5	14.6	-	0	-	-	-	36	-	14.6	14.7	16	0.5	3.4	7.5	36	>6
<a href="#">Lab028</a>	5	14.2	15.0	13	0.9	6.0	17.6	20	2.4	14.2	15.1	102	0.8	5.3	16.9	20	2.6
<a href="#">Lab037</a>	5	12.6	-	0	-	-	-	20	-	12.6	15.7	15	0.5	3.2	31.0	20	-1.4
<a href="#">Lab041</a>	5	12.8	-	0	-	-	-	20	-	12.8	12.9	1	-	-	-	20	-
Total	-	-	14.8	18	0.9	6.1	-	-	-	-	15.1	223	0.8	5.3	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month :

Cumulative : from    to

[TOP](#)

### Control D Month vs. Cumulative

		Control D (Lot No.: AC1203D)															
		Month (2018/08)								CUM (2014/02/01~2018/08/31)							
UnitID <span style="font-size: small;">↑</span>	Reagent Kit (Code) <span style="font-size: small;">↑</span>	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab019</a>	5	3.4	-	0	-	-	-	20	-	3.4	3.6	30	0.6	16.7	39.2	20	0.8
<a href="#">Lab020</a>	5	3.3	-	0	-	-	-	20	-	3.3	3.5	29	0.2	5.7	17.5	20	2.4
<a href="#">Lab021</a>	5	2.8	-	0	-	-	-	20	-	2.8	4.5	12	1.1	24.4	109.6	20	-1.7
<a href="#">Lab022</a>	5	3.3	3.2	5	0.3	9.4	21.8	20	1.8	3.3	3.3	18	0.2	6.1	12.1	20	3.3
<a href="#">Lab026</a>	5	3.2	-	0	-	-	-	36	-	3.2	3.2	16	0.1	3.1	6.3	36	>6
<a href="#">Lab028</a>	5	3.3	3.6	13	0.2	5.6	20.2	20	1.9	3.3	3.6	102	0.2	5.6	20.2	20	1.9
<a href="#">Lab037</a>	5	3.0	-	0	-	-	-	20	-	3.0	3.5	15	0.1	2.9	22.4	20	1.1
<a href="#">Lab041</a>	5	2.9	-	0	-	-	-	20	-	2.9	3.2	1	-	-	-	20	-
Total	-	-	3.5	18	0.3	8.6	-	-	-	-	3.6	223	0.4	11.1	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month :

Cumulative : from    to

[TOP](#)

Reagent Kit	Reagent Code
Medicon	5