

# QC Chart of Internal Quality Control (IQC)for G6PD Quantitative Test

Select LotNo : AH1220N (2019-01-28 ~ 2100-12-31) [Change](#)

[Print Table](#)

## Lab020

	Control N		Control D	
QC Control Lot No.	AH1220N		AH1220D	
Duration of the Analyzing	Month (2019/09)	CUM (2019/02/01~2019/09/30)	Month (2019/09)	CUM (2019/02/01~2019/09/30)
Runs (N)	3	13	3	13
Mean (U/gHb)	15.3	15.2	4.6	4.6
SD	0.1	0.5	0.1	0.2
CV (%)	0.7	3.3	2.2	4.3
Target Value (U/gHb)	15.9	15.9	5.2	5.2
Total Error (%)	5.1	11.0	15.9	20.2
TEa (%)	20	20	20	20
$\sigma$	>6	4.7	3.8	2.0

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 x CV (%)

$\sigma$  (Sigma) = [TEa% - Bias (%) ] / CV (%)


Month : 2019 09 [Change](#) ; Cumulative : from 2019 02 01 to 2019 09 30 [Change](#)

[TOP](#)

## Control N SDI QC Chart

Lot No.: AH1220N ; Duration : 2019-02-01 to 2019-09-30 ; Target : 15.9 ; SD : 0.70

Lab020




Month : 2019 09 [Change](#) ; Cumulative : from 2019 02 01 to 2019 09 30 [Change](#)

[TOP](#)

## Control D SDI QC Chart

Lot No.: AH1220D ; Duration : 2019-02-01 to 2019-09-30 ; Target : 5.2 ; SD : 0.40

Lab020



Month : 2019 09 [Change](#) ; Cumulative : from 2019 02 01 to 2019 09 30 [Change](#)


[TOP](#)

## Control N and Control D SDI QC Chart

Lot No.: AH1220N ; Duration : 2019-02-01 to 2019-09-30 ; Target : 15.9 ; SD : 0.70 ( ● )

Lot No.: AH1220D ; Duration : 2019-02-01 to 2019-09-30 ; Target : 5.2 ; SD : 0.40 ( □ )

Lab020



Date 06-18 07-04 07-11 07-18 07-24 07-30 08-06 08-14 08-20 08-27 09-03 09-10 09-19

Month : 2019 09  ; Cumulative : from 2019 02 01 to 2019 09 30

[\[TOP\]](#)

# Peer Group Statistics (Table 1)

Select LotNo : AH1220N (2019-01-28 ~ 2100-12-31) ▼ Change

Select Reagent Kit : 5 - Medicon ▼ Change

Print Table 1

## Monthly

Month : 2019 ▼ 09 ▼ Change

		Control N (Lot No.: AH1220N)									Control D (Lot No.: AH1220D)						
UnitID ↑	Reagent Kit (Code) ↑	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab020</a>	5	15.9	15.3	3	0.1	0.7	5.1	20	>6	5.2	4.6	3	0.1	2.2	15.9	20	3.8
<a href="#">Lab022</a>	5	12.4	12.7	1	-	-	-	20	-	4.1	3.9	1	-	-	-	20	-
<a href="#">Lab032</a>	5	15.3	14.6	7	0.4	2.7	10.1	20	5.7	5.2	4.4	7	0.1	2.3	19.9	20	2.0
<a href="#">Lab033</a>	5	12.4	12.6	4	0.2	1.6	4.8	20	>6	3.9	3.8	4	0.1	2.6	7.8	20	>6
<b>Total</b>	-	-	14.1	15	1.1	7.8	-	-	-	-	4.2	15	0.3	7.1	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[TOP](#)

## Cumulative

Cumulative : from 2019 ▼ 02 ▼ 01 ▼ to 2019 ▼ 09 ▼ 30 ▼ Change

		Control N (Lot No.: AH1220N)									Control D (Lot No.: AH1220D)						
UnitID ↑	Reagent Kit (Code) ↑	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab020</a>	5	15.9	15.2	13	0.5	3.3	11.0	20	4.7	5.2	4.6	13	0.2	4.3	20.2	20	2.0
<a href="#">Lab022</a>	5	12.4	12.4	18	0.3	2.4	4.8	20	>6	4.1	4.0	18	0.1	2.5	7.4	20	>6
<a href="#">Lab032</a>	5	15.3	14.4	26	0.5	3.5	12.8	20	4.0	5.2	4.6	26	0.5	10.9	33.3	20	0.8
<a href="#">Lab033</a>	5	12.4	12.5	25	0.4	3.2	7.2	20	6.0	3.9	3.9	25	0.1	2.6	5.1	20	>6
<b>Total</b>	-	-	13.5	82	1.2	8.9	-	-	-	-	4.2	82	0.5	11.9	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[TOP](#)

Reagent Kit	Reagent Code
Medicon	5

## Peer Group Statistics (Table 2)

Select LotNo : AH1220N (2019-01-28 ~ 2100-12-31) ▼ Change

Select Reagent Kit : 5 - Medicon ▼ Change

Print Table 2

### Control N Month vs. Cumulative

		Control N (Lot No.: AH1220N)															
		Month (2019/09)									CUM (2019/02/01~2019/09/30)						
UnitID ↑	Reagent Kit (Code) ↑	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab020</a>	5	15.9	15.3	3	0.1	0.7	5.1	20	>6	15.9	15.2	13	0.5	3.3	11.0	20	4.7
<a href="#">Lab022</a>	5	12.4	12.7	1	-	-	-	20	-	12.4	12.4	18	0.3	2.4	4.8	20	>6
<a href="#">Lab032</a>	5	15.3	14.6	7	0.4	2.7	10.1	20	5.7	15.3	14.4	26	0.5	3.5	12.8	20	4.0
<a href="#">Lab033</a>	5	12.4	12.6	4	0.2	1.6	4.8	20	>6	12.4	12.5	25	0.4	3.2	7.2	20	6.0
Total	-	-	14.1	15	1.1	7.8	-	-	-	-	13.5	82	1.2	8.9	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month : 2019 ▼ 09 ▼ Change    Cumulative : from 2019 ▼ 02 ▼ 01 ▼ to 2019 ▼ 09 ▼ 30 ▼ Change

[TOP](#)

### Control D Month vs. Cumulative

		Control D (Lot No.: AH1220D)															
		Month (2019/09)									CUM (2019/02/01~2019/09/30)						
UnitID ↑	Reagent Kit (Code) ↑	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab020</a>	5	5.2	4.6	3	0.1	2.2	15.9	20	3.8	5.2	4.6	13	0.2	4.3	20.2	20	2.0
<a href="#">Lab022</a>	5	4.1	3.9	1	-	-	-	20	-	4.1	4.0	18	0.1	2.5	7.4	20	>6
<a href="#">Lab032</a>	5	5.2	4.4	7	0.1	2.3	19.9	20	2.0	5.2	4.6	26	0.5	10.9	33.3	20	0.8
<a href="#">Lab033</a>	5	3.9	3.8	4	0.1	2.6	7.8	20	>6	3.9	3.9	25	0.1	2.6	5.1	20	>6
Total	-	-	4.2	15	0.3	7.1	-	-	-	-	4.2	82	0.5	11.9	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month : 2019 ▼ 09 ▼ Change    Cumulative : from 2019 ▼ 02 ▼ 01 ▼ to 2019 ▼ 09 ▼ 30 ▼ Change

[TOP](#)

Reagent Kit	Reagent Code
Medicon	5