

QC Chart of Internal Quality Control (IQC)for G6PD Quantitative Test

Select LotNo : AE0909N (2016-01-01 ~ 2100-12-31) [Change](#)

[Print Table](#)

Lab028

QC Control Lot No.	Control N		Control D	
	AE0909N		AE0909D	
Duration of the Analyzing	Month (2020/01)	CUM (2019/12/01~2020/01/31)	Month (2020/01)	CUM (2019/12/01~2020/01/31)
Runs (N)	13	26	13	26
Mean (U/gHb)	13.6	14.1	5.4	5.4
SD	0.7	0.9	0.2	0.2
CV (%)	5.1	6.4	3.7	3.7
Target Value (U/gHb)	14.4	14.4	5.8	5.8
Total Error (%)	15.8	14.8	14.3	14.3
TEa (%)	20	20	20	20
σ	2.8	2.8	3.5	3.5

Bias (%) = [(| Mean - Target |) / Target] x 100%

TE : Total Error(%) = Bias (%) + 2 x CV (%)

σ (Sigma) = [TEa% - Bias (%)] / CV (%)


Month : 2020 01 [Change](#) ; Cumulative : from 2019 12 01 to 2020 01 31 [Change](#)

[\[TOP\]](#)

Control N SDI QC Chart

Lot No.: AE0909N ; Duration : 2019-12-01 to 2020-01-31 ; Target : 14.4 ; SD : 0.72

Lab028



Date 12-03 12-05 12-07 12-10 12-12 12-14 12-17 12-19 12-21 12-24 12-26 12-28 12-31 01-04 01-07 01-09 01-11 01-14 01-15 01-18 01-21 01-23 01-25 01-26 01-28 01-30

Month : 2020 01 [Change](#) ; Cumulative : from 2019 12 01 to 2020 01 31 [Change](#)

[\[TOP\]](#)

Control D SDI QC Chart

Lot No.: AE0909D ; Duration : 2019-12-01 to 2020-01-31 ; Target : 5.8 ; SD : 0.58

Lab028



Date 12-03 12-05 12-07 12-10 12-12 12-14 12-17 12-19 12-21 12-24 12-26 12-28 12-31 01-04 01-07 01-09 01-11 01-14 01-15 01-18 01-21 01-23 01-25 01-26 01-28 01-30

Month : 2020 01 [Change](#) ; Cumulative : from 2019 12 01 to 2020 01 31 [Change](#)


[\[TOP\]](#)

Control N and Control D SDI QC Chart

Lot No.: AE0909N ; Duration : 2019-12-01 to 2020-01-31 ; Target : 14.4 ; SD : 0.72 (●)

Lot No.: AE0909D ; Duration : 2019-12-01 to 2020-01-31 ; Target : 5.8 ; SD : 0.58 (□)

Lab028



Month : 2020 | 01 | ; Cumulative : from 2019 | 12 | 01 to 2020 | 01 | 31 |

[\[TOP\]](#)

Peer Group Statistics (Table 1)

Select LotNo : AE0909N (2016-01-01 ~ 2100-12-31) Change

Select Reagent Kit : 5 - Medicon Change

Print Table 1

Monthly

Month : 2020 01 Change

		Control N (Lot No.: AE0909N)									Control D (Lot No.: AE0909D)						
UnitID	Reagent Kit (Code)	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
Lab019	5	15.3	15.5	5	0.2	1.3	3.9	20	>6	5.3	5.4	5	0.1	1.9	5.6	20	>6
Lab026	5	15.9	12.9	23	0.1	0.8	20.4	20	1.4	5.8	5.0	23	0.1	2.0	17.8	20	3.1
Lab028	5	14.8	13.6	13	0.7	5.1	18.4	20	2.3	4.8	5.4	13	0.2	3.7	19.9	20	2.0
Lab034	5	14.4	14.5	5	1.0	6.9	14.5	20	2.8	5.8	4.7	5	0.2	4.3	27.5	20	0.2
Total	-	-	13.6	46	1.0	7.4	-	-	-	-	5.1	46	0.3	5.9	-	-	-

Bias (%) = [(| Mean - Target |) / Target] x 100%

TE : Total Error(%) = Bias (%) + 2 x CV (%)

σ (Sigma) = [TEa% - Bias (%)] / CV (%)

[TOP](#)

Cumulative

Cumulative : from 2016 02 01 to 2020 01 31 Change

		Control N (Lot No.: AE0909N)									Control D (Lot No.: AE0909D)						
UnitID	Reagent Kit (Code)	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
Lab018	5	14.2	14.9	142	1.2	8.1	21.0	20	1.9	5.5	5.5	142	0.3	5.5	10.9	20	3.6
Lab019	5	15.3	15.2	99	0.3	2.0	4.6	20	>6	5.3	5.3	99	0.1	1.9	3.8	20	>6
Lab020	5	14.4	15.2	28	0.5	3.3	12.1	20	4.4	5.8	5.3	28	0.3	5.7	19.9	20	2.0
Lab026	5	15.9	13.7	318	1.3	9.5	32.8	20	0.6	5.8	5.1	318	0.3	5.9	23.8	20	1.3
Lab027	5	15.5	15.5	107	1.3	8.4	16.8	20	2.4	5.4	5.6	107	0.5	8.9	21.6	20	1.8
Lab028	5	14.8	14.5	134	0.7	4.8	11.7	20	3.7	4.8	5.6	134	0.3	5.4	27.4	20	0.6
Lab032	5	15.3	15.9	38	0.9	5.7	15.2	20	2.8	5.2	5.7	38	0.3	5.3	20.1	20	2.0
Lab033	5	14.4	14.9	49	0.6	4.0	11.5	20	4.1	5.8	5.2	49	0.2	3.8	18.0	20	2.5
Lab034	5	14.4	14.4	78	0.8	5.6	11.1	20	3.6	5.8	5.0	78	0.4	8.0	29.8	20	0.8
Lab037	5	14.9	15.2	30	0.8	5.3	12.5	20	3.4	5.4	5.5	30	0.3	5.5	12.8	20	3.3
Lab040	5	14.4	13.3	39	0.7	5.3	18.2	20	2.3	5.8	4.7	39	0.3	6.4	31.7	20	0.2
Lab043	5	11.5	13.1	84	0.9	6.9	27.7	20	0.9	4.5	4.9	84	0.3	6.1	21.1	20	1.8
Total	-	-	14.4	1146	1.3	9.0	-	-	-	-	5.3	1146	0.4	7.5	-	-	-

Bias (%) = [(| Mean - Target |) / Target] x 100%

TE : Total Error(%) = Bias (%) + 2 x CV (%)

σ (Sigma) = [TEa% - Bias (%)] / CV (%)

[TOP](#)

Reagent Kit	Reagent Code
Medicon	5

Peer Group Statistics (Table 2)

Select LotNo : AE0909N (2016-01-01 ~ 2100-12-31) Change

Select Reagent Kit : 5 - Medicon Change

Print Table 2

Control N Month vs. Cumulative

		Control N (Lot No.: AE0909N)															
		Month (2020/01)								CUM (2016/02/01~2020/01/31)							
UnitID	Reagent Kit (Code)	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
Lab018	5	14.2	-	0	-	-	-	20	-	14.2	14.9	142	1.2	8.1	21.0	20	1.9
Lab019	5	15.3	15.5	5	0.2	1.3	3.9	20	>6	15.3	15.2	99	0.3	2.0	4.6	20	>6
Lab020	5	14.4	-	0	-	-	-	20	-	14.4	15.2	28	0.5	3.3	12.1	20	4.4
Lab026	5	15.9	12.9	23	0.1	0.8	20.4	20	1.4	15.9	13.7	318	1.3	9.5	32.8	20	0.6
Lab027	5	15.5	-	0	-	-	-	20	-	15.5	15.5	107	1.3	8.4	16.8	20	2.4
Lab028	5	14.8	13.6	13	0.7	5.1	18.4	20	2.3	14.8	14.5	134	0.7	4.8	11.7	20	3.7
Lab032	5	15.3	-	0	-	-	-	20	-	15.3	15.9	38	0.9	5.7	15.2	20	2.8
Lab033	5	14.4	-	0	-	-	-	20	-	14.4	14.9	49	0.6	4.0	11.5	20	4.1
Lab034	5	14.4	14.5	5	1.0	6.9	14.5	20	2.8	14.4	14.4	78	0.8	5.6	11.1	20	3.6
Lab037	5	14.9	-	0	-	-	-	20	-	14.9	15.2	30	0.8	5.3	12.5	20	3.4
Lab040	5	14.4	-	0	-	-	-	20	-	14.4	13.3	39	0.7	5.3	18.2	20	2.3
Lab043	5	11.5	-	0	-	-	-	20	-	11.5	13.1	84	0.9	6.9	27.7	20	0.9
Total	-	-	13.6	46	1.0	7.4	-	-	-	-	14.4	1146	1.3	9.0	-	-	-

Bias (%) = [(| Mean - Target |) / Target] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%)] / CV (%)

Month : 2020 01 Change

Cumulative : from 2016 02 01 to 2020 01 31 Change

[TOP](#)

Control D Month vs. Cumulative

		Control D (Lot No.: AE0909D)															
		Month (2020/01)								CUM (2016/02/01~2020/01/31)							
UnitID	Reagent Kit (Code)	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
Lab018	5	5.5	-	0	-	-	-	20	-	5.5	5.5	142	0.3	5.5	10.9	20	3.6
Lab019	5	5.3	5.4	5	0.1	1.9	5.6	20	>6	5.3	5.3	99	0.1	1.9	3.8	20	>6
Lab020	5	5.8	-	0	-	-	-	20	-	5.8	5.3	28	0.3	5.7	19.9	20	2.0
Lab026	5	5.8	5.0	23	0.1	2.0	17.8	20	3.1	5.8	5.1	318	0.3	5.9	23.8	20	1.3
Lab027	5	5.4	-	0	-	-	-	20	-	5.4	5.6	107	0.5	8.9	21.6	20	1.8
Lab028	5	4.8	5.4	13	0.2	3.7	19.9	20	2.0	4.8	5.6	134	0.3	5.4	27.4	20	0.6
Lab032	5	5.2	-	0	-	-	-	20	-	5.2	5.7	38	0.3	5.3	20.1	20	2.0
Lab033	5	5.8	-	0	-	-	-	20	-	5.8	5.2	49	0.2	3.8	18.0	20	2.5
Lab034	5	5.8	4.7	5	0.2	4.3	27.5	20	0.2	5.8	5.0	78	0.4	8.0	29.8	20	0.8
Lab037	5	5.4	-	0	-	-	-	20	-	5.4	5.5	30	0.3	5.5	12.8	20	3.3
Lab040	5	5.8	-	0	-	-	-	20	-	5.8	4.7	39	0.3	6.4	31.7	20	0.2
Lab043	5	4.5	-	0	-	-	-	20	-	4.5	4.9	84	0.3	6.1	21.1	20	1.8
Total	-	-	5.1	46	0.3	5.9	-	-	-	-	5.3	1146	0.4	7.5	-	-	-

Bias (%) = [(| Mean - Target |) / Target] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%)] / CV (%)

Month : 2020 01 Change

Cumulative : from 2016 02 01 to 2020 01 31 Change

[TOP](#)

Reagent Kit	Reagent Code
Medicon	5