

# G6PD 定量檢驗內部品管統計結果及 SDI 分佈圖

選擇 LotNo : BJ0922N (2021-01-01 ~ 2100-12-31) Change

列印圖表

檢驗單位 : **Lab014**

QC Control Lot No.	Control N		Control D	
	BJ0922N		BJ0922D	
檢驗期間	Month (2022/01)	CUM (2021/10/29~2022/01/29)	Month (2022/01)	CUM (2021/10/29~2022/01/29)
累計點數 (N)	7	26	7	26
累計平均值 (Mean) U/gHb	14.4	14.5	5.5	5.6
累計標準差 (SD)	0.3	0.4	0.1	0.2
累計變異係數 (CV) %	2.1	2.8	1.8	3.6
目標值 (Target) U/gHb	14.6	14.6	5.5	5.5
總誤差 (TE) %	5.5	6.2	3.6	9.0
可容許總誤差 (TEa) %	20	20	20	20
$\sigma$	>6	>6	>6	5.1

誤差 (Bias)% =  $\{ [ | \text{累計平均值 (Mean)} - \text{目標值 (Target)} | ] / \text{目標值 (Target)} \} \times 100\%$   
 總誤差 (TE, Total Error)% = 誤差 (Bias)% + 2 × 累計變異係數 (CV)  
 $\sigma$  (Sigma) = (TEa% - 誤差 (Bias)% ) / 累計變異係數 (CV)

月統計 : 2022 01 Change ; 累計期間 (CUM) : 自 2021 10 29 至 2022 01 29 Change


[TOP](#)

## Control N 及 Control D SDI 分佈圖

Lot No.: BJ0922N; 檢驗期間: 2021-10-29 至 2022-01-29; 目標值 (Target): 14.6; 參考標準差 (SD): 0.40 (●)

Lot No.: BJ0922D; 檢驗期間: 2021-10-29 至 2022-01-29; 目標值 (Target): 5.5; 參考標準差 (SD): 0.20 (□)

單位: Lab014




月統計 : 2022 01 Change ; 累計期間 (CUM) : 自 2021 10 29 至 2022 01 29 Change

[TOP](#)

## Control N SDI 分佈圖

Lot No.: BJ0922N; 檢驗期間: 2021-10-29 至 2022-01-29; 目標值 (Target): 14.6; 參考標準差 (SD): 0.40

單位: Lab014




月統計 : 2022 01 Change ; 累計期間 (CUM) : 自 2021 10 29 至 2022 01 29 Change

[TOP](#)

## Control D SDI 分佈圖

Lot No.: BJ0922D; 檢驗期間: 2021-10-29 至 2022-01-29; 目標值 (Target): 5.5; 參考標準差 (SD): 0.20

單位: Lab014



月統計: 2022 01 Change ; 累計期間 (CUM): 自 2021 10 29 至 2022 01 29 Change

[\[TOP\]](#)

# 各檢驗單位 G6PD 定量檢驗 內部品管統計結果 (總表一)

選擇 LotNo : BJ0922N (2021-01-01 ~ 2100-12-31) Change

選擇 Reagent Kit : 3 - Innovation Change

[列印總表一](#)

## 月統計表

統計月份 : 2022 01 Change

UnitID	Reagent Kit (Code)	Control N (Lot No.: BJ0922N)								Control D (Lot No.: BJ0922D)							
		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	$\sigma$	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	$\sigma$
<a href="#">Lab004</a>	3	16.2	16.6	4	0.2	1.2	4.9	20	>6	5.2	5.2	4	0.2	3.8	7.7	20	5.3
<a href="#">Lab014</a>	3	14.6	14.4	7	0.3	2.1	5.5	20	>6	5.5	5.5	7	0.1	1.8	3.6	20	>6
<a href="#">Lab016</a>	3	13.8	15.0	4	0.2	1.3	11.4	20	>6	4.7	4.8	4	0.1	2.1	6.3	20	>6
<a href="#">Lab051</a>	3	16.0	14.5	4	0.2	1.4	12.1	20	>6	5.4	5.6	4	0.1	1.8	7.3	20	>6
Total	-	-	15.0	19	0.9	6.0	-	-	-	-	5.3	19	0.3	5.7	-	-	-

誤差 (Bias)% =  $\{ [ | \text{累計平均值 (Mean)} - \text{目標值 (Target)} | ] / \text{目標值 (Target)} \} \times 100\%$

總誤差 (TE, Total Error)% = 誤差 (Bias)% + 2 × 累計變異係數 (CV)

$\sigma$  (Sigma) = (TEa% - 誤差 (Bias)% ) / 累計變異係數 (CV)

[TOP](#)

## 累計統計表

累計期間 : 自 2021 01 01 至 2022 01 31 Change

UnitID	Reagent Kit (Code)	Control N (Lot No.: BJ0922N)								Control D (Lot No.: BJ0922D)							
		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	$\sigma$	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	$\sigma$
<a href="#">Lab004</a>	3	16.2	16.3	44	0.4	2.5	5.5	20	>6	5.2	5.2	44	0.2	3.8	7.7	20	5.3
<a href="#">Lab013</a>	3	17.1	16.9	41	0.2	1.2	3.5	20	>6	5.6	5.6	41	0.2	3.6	7.1	20	5.6
<a href="#">Lab014</a>	3	14.6	14.5	47	0.4	2.8	6.2	20	>6	5.5	5.6	47	0.2	3.6	9.0	20	5.1
<a href="#">Lab016</a>	3	13.8	14.8	48	0.4	2.7	12.7	20	4.7	4.7	4.8	48	0.2	4.2	10.5	20	4.3
<a href="#">Lab051</a>	3	16.0	14.5	6	0.2	1.4	12.1	20	>6	5.4	5.6	6	0.1	1.8	7.3	20	>6
Total	-	-	15.5	186	1.1	7.1	-	-	-	-	5.3	186	0.4	7.5	-	-	-

誤差 (Bias)% =  $\{ [ | \text{累計平均值 (Mean)} - \text{目標值 (Target)} | ] / \text{目標值 (Target)} \} \times 100\%$

總誤差 (TE, Total Error)% = 誤差 (Bias)% + 2 × 累計變異係數 (CV)

$\sigma$  (Sigma) = (TEa% - 誤差 (Bias)% ) / 累計變異係數 (CV)

[TOP](#)

Reagent Kit	Reagent Code
Innovation	3

# 各檢驗單位 G6PD 定量檢驗 內部品管統計結果 (總表二)

選擇 LotNo :

選擇 Reagent Kit :

[列印總表二](#)

## Control N 月統計與累計統計之比較表

		Control N (Lot No.: BJ0922N)															
		Month (2022/01)								CUM (2021/01/01~2022/01/31)							
UnitID <input type="button" value="↑"/>	Reagent Kit (Code) <input type="button" value="↑"/>	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab004</a>	3	16.2	16.6	4	0.2	1.2	4.9	20	>6	16.2	16.3	44	0.4	2.5	5.5	20	>6
<a href="#">Lab013</a>	3	17.1	-	0	-	-	-	20	-	17.1	16.9	41	0.2	1.2	3.5	20	>6
<a href="#">Lab014</a>	3	14.6	14.4	7	0.3	2.1	5.5	20	>6	14.6	14.5	47	0.4	2.8	6.2	20	>6
<a href="#">Lab016</a>	3	13.8	15.0	4	0.2	1.3	11.4	20	>6	13.8	14.8	48	0.4	2.7	12.7	20	4.7
<a href="#">Lab051</a>	3	16.0	14.5	4	0.2	1.4	12.1	20	>6	16.0	14.5	6	0.2	1.4	12.1	20	>6
Total	-	-	15.0	19	0.9	6.0	-	-	-	-	15.5	186	1.1	7.1	-	-	-

誤差 (Bias)% = { [ | 累計平均值 (Mean) - 目標值 (Target) | ] / 目標值 (Target) } x 100%

總誤差 (TE, Total Error)% = 誤差 (Bias)% + 2 × 累計變異係數 (CV)

σ (Sigma) = (TEa% - 誤差 (Bias)% ) / 累計變異係數 (CV)

月統計 :

累計期間 (CUM) : 自    至

[TOP](#)

## Control D 月統計與累計統計之比較表

		Control D (Lot No.: BJ0922D)															
		Month (2022/01)								CUM (2021/01/01~2022/01/31)							
UnitID <input type="button" value="↑"/>	Reagent Kit (Code) <input type="button" value="↑"/>	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab004</a>	3	5.2	5.2	4	0.2	3.8	7.7	20	5.3	5.2	5.2	44	0.2	3.8	7.7	20	5.3
<a href="#">Lab013</a>	3	5.6	-	0	-	-	-	20	-	5.6	5.6	41	0.2	3.6	7.1	20	5.6
<a href="#">Lab014</a>	3	5.5	5.5	7	0.1	1.8	3.6	20	>6	5.5	5.6	47	0.2	3.6	9.0	20	5.1
<a href="#">Lab016</a>	3	4.7	4.8	4	0.1	2.1	6.3	20	>6	4.7	4.8	48	0.2	4.2	10.5	20	4.3
<a href="#">Lab051</a>	3	5.4	5.6	4	0.1	1.8	7.3	20	>6	5.4	5.6	6	0.1	1.8	7.3	20	>6
Total	-	-	5.3	19	0.3	5.7	-	-	-	-	5.3	186	0.4	7.5	-	-	-

誤差 (Bias)% = { [ | 累計平均值 (Mean) - 目標值 (Target) | ] / 目標值 (Target) } x 100%

總誤差 (TE, Total Error)% = 誤差 (Bias)% + 2 × 累計變異係數 (CV)

σ (Sigma) = (TEa% - 誤差 (Bias)% ) / 累計變異係數 (CV)

月統計 :

累計期間 (CUM) : 自    至

[TOP](#)

Reagent Kit	Reagent Code
Innovation	3