


# Summary Report of IQC program for G6PD Quantitative Test - Medicon Group - August 2022 -

## I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:BJ0922N)	Control D (Lot No.: BJ0922D)
Labs	8	8
Received results number (n)	33	33
Median	14.3 (U/g Hb)	4.6 (U/g Hb)
Mean	14.2 (U/g Hb)	4.6 (U/g Hb)
SD	0.5	0.2
CV	3.5%	4.3%
Range of G6PD	13.0 ~ 15.0 (U/g Hb)	4.3 ~ 4.9 (U/g Hb)
Range of Hb	2.0 ~ 3.1 (g/dL)	1.9 ~ 2.7 (g/dL)

\*The statistic results are calculated from all labs reported in this month  
 \*\* G6PD Method = Medicon reagent kit, 37°C

## II. The distribution of G6PD reported for each lab in this survey



# QC Chart of Internal Quality Control (IQC)for G6PD Quantitative Test

Select LotNo : BJ0922N (2021-01-01 ~ 2100-12-31) Change

[Print Table](#)

## Lab043

QC Control Lot No.	Control N		Control D	
	BJ0922N		BJ0922D	
Duration of the Analyzing	Month (2022/08)	CUM (2022/02/10~2022/08/31)	Month (2022/08)	CUM (2022/02/10~2022/08/31)
Runs (N)	4	26	4	26
Mean (U/gHb)	14.6	14.5	4.8	4.7
SD	0.1	1.0	0.1	0.3
CV (%)	0.7	6.9	2.1	6.4
Target Value (U/gHb)	13.2	13.2	4.2	4.2
Total Error (%)	12.0	23.6	18.5	24.7
TEa (%)	20	20	20	20
$\sigma$	>6	1.5	2.7	1.3

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%  
 TE : Total Error(%) = Bias (%) + 2 x CV (%)  
 $\sigma$  (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month : 2022 08 Change ; Cumulative : from 2022 02 10 to 2022 08 31 Change


[TOP](#)

## Control N and Control D SDI QC Chart

Lot No.: BJ0922N ; Duration : 2022-02-10 to 2022-08-31 ; Target : 13.2 ; SD : 1.50 (●)

Lot No.: BJ0922D ; Duration : 2022-02-10 to 2022-08-31 ; Target : 4.2 ; SD : 0.50 (□)

Lab043



Date 02-10 02-12 02-15 02-17 02-19 02-24 04-06 04-28 04-29 05-05 05-19 05-26 06-02 06-06 06-09 06-16 06-23 06-30 07-07 07-14 07-21 07-28 08-04 08-11 08-18 08-25


Month : 2022 08 Change ; Cumulative : from 2022 02 10 to 2022 08 31 Change

[TOP](#)

## Control N SDI QC Chart

Lot No.: BJ0922N ; Duration : 2022-02-10 to 2022-08-31 ; Target : 13.2 ; SD : 1.50

Lab043



Date 02-10 02-12 02-15 02-17 02-19 02-24 04-06 04-28 04-29 05-05 05-19 05-26 06-02 06-06 06-09 06-16 06-23 06-30 07-07 07-14 07-21 07-28 08-04 08-11 08-18 08-25


Month : 2022 08 Change ; Cumulative : from 2022 02 10 to 2022 08 31 Change

[TOP](#)

## Control D SDI QC Chart

Lot No.: BJ0922D ; Duration : 2022-02-10 to 2022-08-31 ; Target : 4.2 ; SD : 0.50

Lab043



Month :    ; Cumulative : from    to

[TOP](#)

## Peer Group Statistics (Table 1)

Select LotNo : BJ0922N (2021-01-01 ~ 2100-12-31) Change

Select Reagent Kit : 5 - Medicon Change

Print Table 1

### Monthly

Month : 2022 08 Change

UnitID <small>↑</small>	Reagent Kit (Code) <small>↑</small>	Control N (Lot No.: BJ0922N)								Control D (Lot No.: BJ0922D)							
		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab019</a>	5	14.6	13.9	4	0.1	0.7	6.2	20	>6	4.6	4.5	4	0.2	4.4	11.1	20	4.1
<a href="#">Lab020</a>	5	14.4	14.6	4	0.3	2.1	5.5	20	>6	4.6	4.8	4	0.1	2.1	8.5	20	>6
<a href="#">Lab032</a>	5	14.9	14.3	4	0.3	2.1	8.2	20	>6	4.8	4.5	4	0.1	2.2	10.7	20	>6
<a href="#">Lab033</a>	5	13.8	13.8	3	0.5	3.6	7.2	20	5.6	4.7	4.7	3	0.2	4.3	8.5	20	4.7
<a href="#">Lab037</a>	5	13.8	14.5	5	0.2	1.4	7.8	20	>6	4.7	4.6	5	0.2	4.3	10.8	20	4.2
<a href="#">Lab040</a>	5	14.4	14.7	4	0.2	1.4	4.8	20	>6	4.6	4.5	4	0.1	2.2	6.6	20	>6
<a href="#">Lab043</a>	5	13.2	14.6	4	0.1	0.7	12.0	20	>6	4.2	4.8	4	0.1	2.1	18.5	20	2.7
<a href="#">Lab049</a>	5	13.8	13.4	5	0.3	2.2	7.4	20	>6	4.7	4.5	5	0.2	4.4	13.1	20	3.6
Total	-	-	14.2	33	0.5	3.5	-	-	-	-	4.6	33	0.2	4.3	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[TOP](#)

### Cumulative

Cumulative : from 2021 01 01 to 2022 08 31 Change

UnitID <small>↑</small>	Reagent Kit (Code) <small>↑</small>	Control N (Lot No.: BJ0922N)								Control D (Lot No.: BJ0922D)							
		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab018</a>	5	16.2	16.4	8	0.4	2.4	6.1	20	>6	5.5	5.5	8	0.2	3.6	7.3	20	5.6
<a href="#">Lab019</a>	5	14.6	14.4	32	0.6	4.2	9.7	20	4.4	4.6	4.6	32	0.1	2.2	4.3	20	>6
<a href="#">Lab020</a>	5	14.4	14.5	40	0.6	4.1	9.0	20	4.7	4.6	4.7	40	0.2	4.3	10.7	20	4.1
<a href="#">Lab022</a>	5	14.4	14.3	12	0.6	4.2	9.1	20	4.6	4.6	4.6	12	0.3	6.5	13.0	20	3.1
<a href="#">Lab028</a>	5	15.2	15.1	34	1.1	7.3	15.2	20	2.6	4.9	4.9	34	0.3	6.1	12.2	20	3.3
<a href="#">Lab032</a>	5	14.9	14.4	88	0.4	2.8	8.9	20	5.9	4.8	4.5	88	0.1	2.2	10.7	20	>6
<a href="#">Lab033</a>	5	13.8	13.5	33	0.6	4.4	11.1	20	4.1	4.7	4.5	33	0.3	6.7	17.6	20	2.3
<a href="#">Lab037</a>	5	13.8	14.7	38	0.4	2.7	12.0	20	5.0	4.7	4.7	38	0.2	4.3	8.5	20	4.7
<a href="#">Lab040</a>	5	14.4	14.4	9	0.3	2.1	4.2	20	>6	4.6	4.5	9	0.1	2.2	6.6	20	>6
<a href="#">Lab043</a>	5	13.2	13.4	55	1.5	11.2	23.9	20	1.7	4.2	4.3	55	0.5	11.6	25.6	20	1.5
<a href="#">Lab049</a>	5	13.8	14.7	65	0.7	4.8	16.0	20	2.8	4.7	4.7	65	0.2	4.3	8.5	20	4.7
Total	-	-	14.4	414	1.0	6.9	-	-	-	-	4.6	414	0.3	6.5	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[TOP](#)

Reagent Kit	Reagent Code
Medicon	5

## Peer Group Statistics (Table 2)

Select LotNo : BJ0922N (2021-01-01 ~ 2100-12-31) Change

Select Reagent Kit : 5 - Medicon Change

[Print Table 2](#)

### Control N Month vs. Cumulative

Control N (Lot No.: BJ0922N)																	
Month (2022/08)										CUM (2021/01/01~2022/08/31)							
UnitID ↑	Reagent Kit (Code) ↑	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab018</a>	5	16.2	-	0	-	-	-	20	-	16.2	16.4	8	0.4	2.4	6.1	20	>6
<a href="#">Lab019</a>	5	14.6	13.9	4	0.1	0.7	6.2	20	>6	14.6	14.4	32	0.6	4.2	9.7	20	4.4
<a href="#">Lab020</a>	5	14.4	14.6	4	0.3	2.1	5.5	20	>6	14.4	14.5	40	0.6	4.1	9.0	20	4.7
<a href="#">Lab022</a>	5	14.4	-	0	-	-	-	20	-	14.4	14.3	12	0.6	4.2	9.1	20	4.6
<a href="#">Lab028</a>	5	15.2	-	0	-	-	-	20	-	15.2	15.1	34	1.1	7.3	15.2	20	2.6
<a href="#">Lab032</a>	5	14.9	14.3	4	0.3	2.1	8.2	20	>6	14.9	14.4	88	0.4	2.8	8.9	20	5.9
<a href="#">Lab033</a>	5	13.8	13.8	3	0.5	3.6	7.2	20	5.6	13.8	13.5	33	0.6	4.4	11.1	20	4.1
<a href="#">Lab037</a>	5	13.8	14.5	5	0.2	1.4	7.8	20	>6	13.8	14.7	38	0.4	2.7	12.0	20	5.0
<a href="#">Lab040</a>	5	14.4	14.7	4	0.2	1.4	4.8	20	>6	14.4	14.4	9	0.3	2.1	4.2	20	>6
<a href="#">Lab043</a>	5	13.2	14.6	4	0.1	0.7	12.0	20	>6	13.2	13.4	55	1.5	11.2	23.9	20	1.7
<a href="#">Lab049</a>	5	13.8	13.4	5	0.3	2.2	7.4	20	>6	13.8	14.7	65	0.7	4.8	16.0	20	2.8
Total	-	-	14.2	33	0.5	3.5	-	-	-	-	14.4	414	1.0	6.9	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month : 2022 08 Change

Cumulative : from 2021 01 01 to 2022 08 31 Change

[TOP](#)

### Control D Month vs. Cumulative

Control D (Lot No.: BJ0922D)																	
Month (2022/08)										CUM (2021/01/01~2022/08/31)							
UnitID ↑	Reagent Kit (Code) ↑	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab018</a>	5	5.5	-	0	-	-	-	20	-	5.5	5.5	8	0.2	3.6	7.3	20	5.6
<a href="#">Lab019</a>	5	4.6	4.5	4	0.2	4.4	11.1	20	4.1	4.6	4.6	32	0.1	2.2	4.3	20	>6
<a href="#">Lab020</a>	5	4.6	4.8	4	0.1	2.1	8.5	20	>6	4.6	4.7	40	0.2	4.3	10.7	20	4.1
<a href="#">Lab022</a>	5	4.6	-	0	-	-	-	20	-	4.6	4.6	12	0.3	6.5	13.0	20	3.1
<a href="#">Lab028</a>	5	4.9	-	0	-	-	-	20	-	4.9	4.9	34	0.3	6.1	12.2	20	3.3
<a href="#">Lab032</a>	5	4.8	4.5	4	0.1	2.2	10.7	20	>6	4.8	4.5	88	0.1	2.2	10.7	20	>6
<a href="#">Lab033</a>	5	4.7	4.7	3	0.2	4.3	8.5	20	4.7	4.7	4.5	33	0.3	6.7	17.6	20	2.3
<a href="#">Lab037</a>	5	4.7	4.6	5	0.2	4.3	10.8	20	4.2	4.7	4.7	38	0.2	4.3	8.5	20	4.7
<a href="#">Lab040</a>	5	4.6	4.5	4	0.1	2.2	6.6	20	>6	4.6	4.5	9	0.1	2.2	6.6	20	>6
<a href="#">Lab043</a>	5	4.2	4.8	4	0.1	2.1	18.5	20	2.7	4.2	4.3	55	0.5	11.6	25.6	20	1.5
<a href="#">Lab049</a>	5	4.7	4.5	5	0.2	4.4	13.1	20	3.6	4.7	4.7	65	0.2	4.3	8.5	20	4.7
Total	-	-	4.6	33	0.2	4.3	-	-	-	-	4.6	414	0.3	6.5	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month : 2022 08 Change

Cumulative : from 2021 01 01 to 2022 08 31 Change

[TOP](#)

Reagent Kit	Reagent Code
Medicon	5