

QC Chart of Internal Quality Control (IQC)for G6PD Quantitative Test

Select LotNo : BJ0922N (2021-01-01 ~ 2100-12-31) Change

Print Table

Lab043

QC Control Lot No.	Control N		Control D	
	BJ0922N		BJ0922D	
	Month (2025/12)	CUM (2025/06/19~2025/12/25)	Month (2025/12)	CUM (2025/06/19~2025/12/25)
Runs (N)	4	26	4	26
Mean (U/gHb)	13.5	14.3	4.7	4.8
SD	0.2	0.6	0.1	0.1
CV (%)	1.5	4.2	2.1	2.1
Target Value (U/gHb)	13.2	13.2	4.2	4.2
Total Error (%)	5.2	16.7	16.2	18.5
TEa (%)	20	20	20	20
σ	>6	2.8	3.9	2.7

Bias (%) = [(| Mean - Target |) / Target] x 100%
TE : Total Error(%) = Bias (%) + 2 × CV (%)
σ (Sigma) = [TEa% - Bias (%)] / CV (%)

Month : 2025 12 Change ; Cumulative : from 2025 06 19 to 2025 12 25 Change


TOP

Control N and Control D SDI
QC Chart

Lot No.: BJ0922N ; Duration : 2025-06-19 to 2025-12-25 ; Target : 13.2 ; SD : 1.50 (●)

Lot No.: BJ0922D ; Duration : 2025-06-19 to 2025-12-25 ; Target : 4.2 ; SD : 0.50 (□)

Lab043




TOP

Control N SDI QC Chart

Lot No.: BJ0922N ; Duration : 2025-06-19 to 2025-12-25 ; Target : 13.2 ; SD : 1.50

Lab043




TOP

Control D SDI QC Chart

Lot No.: BJ0922D ; Duration : 2025-06-19 to 2025-12-25 ; Target : 4.2 ; SD : 0.50

Lab043



Date 06-19 06-26 07-03 07-10 07-17 07-24 07-31 08-07 08-14 08-21 08-28 09-04 09-11 09-18 09-25 10-02 10-09 10-16 10-23 10-30 11-06 11-13 11-20 11-27 12-04 12-11 12-18 12-25

Month : 2025 12 Change ; Cumulative : from 2025 06 19 to 2025 12 25 Change

[TOP](#)

Peer Group Statistics (Table 1)

Select LotNo : BJ0922N (2021-01-01 ~ 2100-12-31) Change

Select Reagent Kit : 5 - Medicon Change

[Print Table 1](#)

Monthly

Month : 2025 12 Change

		Control N (Lot No.: BJ0922N)									Control D (Lot No.: BJ0922D)							
UnitID	Reagent Kit (Code)	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
Lab034	5	13.8	15.2	2	-	-	-	20	-		4.7	4.3	2	-	-	-	20	-
Lab043	5	13.2	13.5	4	0.2	1.5	5.2	20	>6		4.2	4.7	4	0.1	2.1	16.2	20	3.9
Lab049	5	13.8	13.3	4	0.5	3.8	11.1	20	4.3		4.7	4.9	4	0.1	2.0	8.3	20	>6
Lab050	5	13.8	14.2	1	-	-	-	20	-		4.7	5.5	1	-	-	-	20	-
Total	-	-	13.8	11	0.8	5.8	-	-	-		-	4.8	11	0.3	6.3	-	-	-

Bias (%) = [(| Mean - Target |) / Target] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%)] / CV (%)

[\[TOP\]](#)

Cumulative

Cumulative : from 2021 01 01 to 2025 12 31 Change

		Control N (Lot No.: BJ0922N)									Control D (Lot No.: BJ0922D)							
UnitID	Reagent Kit (Code)	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
Lab018	5	16.2	15.3	113	0.4	2.6	10.8	20	5.6		5.5	5.4	113	0.1	1.9	5.5	20	>6
Lab019	5	14.7	14.6	177	0.5	3.4	7.5	20	5.7		5.0	4.8	177	0.2	4.2	12.3	20	3.8
Lab020	5	14.4	14.4	90	0.5	3.5	6.9	20	5.7		4.6	4.7	90	0.2	4.3	10.7	20	4.1
Lab022	5	14.4	14.4	27	0.5	3.5	6.9	20	5.7		4.6	4.7	27	0.3	6.4	14.9	20	2.8
Lab025	5	14.4	14.2	13	0.5	3.5	8.4	20	5.3		5.4	5.1	13	0.3	5.9	17.3	20	2.4
Lab028	5	15.2	15.0	66	0.9	6.0	13.3	20	3.1		4.9	5.1	66	0.3	5.9	15.8	20	2.7
Lab032	5	14.9	14.3	153	0.4	2.8	9.6	20	5.7		4.8	4.5	153	0.2	4.4	15.1	20	3.1
Lab033	5	13.8	13.6	81	0.7	5.1	11.7	20	3.6		4.7	4.6	81	0.3	6.5	15.2	20	2.7
Lab034	5	13.8	14.5	53	0.6	4.1	13.3	20	3.6		4.7	5.0	53	0.7	14.0	34.4	20	1.0
Lab037	5	13.8	14.5	59	0.5	3.4	12.0	20	4.4		4.7	4.7	59	0.2	4.3	8.5	20	4.7
Lab040	5	14.4	14.8	169	0.2	1.4	5.5	20	>6		4.6	4.6	169	0.1	2.2	4.3	20	>6
Lab043	5	13.2	14.1	220	0.9	6.4	19.6	20	2.1		4.2	4.7	220	0.3	6.4	24.7	20	1.3
Lab049	5	13.8	14.3	289	0.6	4.2	12.0	20	3.9		4.7	4.7	289	0.3	6.4	12.8	20	3.1
Lab050	5	13.8	14.4	117	0.6	4.2	12.7	20	3.7		4.7	5.1	117	0.5	9.8	28.1	20	1.2
Total	-	-	14.5	1627	0.7	4.8	-	-	-		-	4.8	1627	0.4	8.3	-	-	-

Bias (%) = [(| Mean - Target |) / Target] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%)] / CV (%)

[\[TOP\]](#)

Reagent Kit	Reagent Code
Medicon	5

Peer Group Statistics (Table 2)

Select LotNo : BJ0922N (2021-01-01 ~ 2100-12-31) Change

Select Reagent Kit : 5 - Medicon Change

Print Table 2

Control N Month vs. Cumulative

		Control N (Lot No.: BJ0922N)															
		Month (2025/12)								CUM (2021/01/01~2025/12/31)							
UnitID ↑	Reagent Kit (Code) ↑	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
Lab018	5	16.2	-	0	-	-	-	20	-	16.2	15.3	113	0.4	2.6	10.8	20	5.6
Lab019	5	14.7	-	0	-	-	-	20	-	14.7	14.6	177	0.5	3.4	7.5	20	5.7
Lab020	5	14.4	-	0	-	-	-	20	-	14.4	14.4	90	0.5	3.5	6.9	20	5.7
Lab022	5	14.4	-	0	-	-	-	20	-	14.4	14.4	27	0.5	3.5	6.9	20	5.7
Lab025	5	14.4	-	0	-	-	-	20	-	14.4	14.2	13	0.5	3.5	8.4	20	5.3
Lab028	5	15.2	-	0	-	-	-	20	-	15.2	15.0	66	0.9	6.0	13.3	20	3.1
Lab032	5	14.9	-	0	-	-	-	20	-	14.9	14.3	153	0.4	2.8	9.6	20	5.7
Lab033	5	13.8	-	0	-	-	-	20	-	13.8	13.6	81	0.7	5.1	11.7	20	3.6
Lab034	5	13.8	15.2	2	-	-	-	20	-	13.8	14.5	53	0.6	4.1	13.3	20	3.6
Lab037	5	13.8	-	0	-	-	-	20	-	13.8	14.5	59	0.5	3.4	12.0	20	4.4
Lab040	5	14.4	-	0	-	-	-	20	-	14.4	14.8	169	0.2	1.4	5.5	20	>6
Lab043	5	13.2	13.5	4	0.2	1.5	5.2	20	>6	13.2	14.1	220	0.9	6.4	19.6	20	2.1
Lab049	5	13.8	13.3	4	0.5	3.8	11.1	20	4.3	13.8	14.3	289	0.6	4.2	12.0	20	3.9
Lab050	5	13.8	14.2	1	-	-	-	20	-	13.8	14.4	117	0.6	4.2	12.7	20	3.7
Total	-	-	13.8	11	0.8	5.8	-	-	-	-	14.5	1627	0.7	4.8	-	-	-

Bias (%) = [(| Mean - Target |) / Target] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%)] / CV (%)

Month : 2025 12 Change

Cumulative : from 2021 01 01 to 2025 12 31 Change

TOP

Control D Month vs. Cumulative

		Control D (Lot No.: BJ0922D)															
		Month (2025/12)								CUM (2021/01/01~2025/12/31)							
UnitID ↑	Reagent Kit (Code) ↑	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
Lab018	5	5.5	-	0	-	-	-	20	-	5.5	5.4	113	0.1	1.9	5.5	20	>6
Lab019	5	5.0	-	0	-	-	-	20	-	5.0	4.8	177	0.2	4.2	12.3	20	3.8
Lab020	5	4.6	-	0	-	-	-	20	-	4.6	4.7	90	0.2	4.3	10.7	20	4.1
Lab022	5	4.6	-	0	-	-	-	20	-	4.6	4.7	27	0.3	6.4	14.9	20	2.8
Lab025	5	5.4	-	0	-	-	-	20	-	5.4	5.1	13	0.3	5.9	17.3	20	2.4
Lab028	5	4.9	-	0	-	-	-	20	-	4.9	5.1	66	0.3	5.9	15.8	20	2.7
Lab032	5	4.8	-	0	-	-	-	20	-	4.8	4.5	153	0.2	4.4	15.1	20	3.1
Lab033	5	4.7	-	0	-	-	-	20	-	4.7	4.6	81	0.3	6.5	15.2	20	2.7
Lab034	5	4.7	4.3	2	-	-	-	20	-	4.7	5.0	53	0.7	14.0	34.4	20	1.0
Lab037	5	4.7	-	0	-	-	-	20	-	4.7	4.7	59	0.2	4.3	8.5	20	4.7
Lab040	5	4.6	-	0	-	-	-	20	-	4.6	4.6	169	0.1	2.2	4.3	20	>6
Lab043	5	4.2	4.7	4	0.1	2.1	16.2	20	3.9	4.2	4.7	220	0.3	6.4	24.7	20	1.3
Lab049	5	4.7	4.9	4	0.1	2.0	8.3	20	>6	4.7	4.7	289	0.3	6.4	12.8	20	3.1
Lab050	5	4.7	5.5	1	-	-	-	20	-	4.7	5.1	117	0.5	9.8	28.1	20	1.2
Total	-	-	4.8	11	0.3	6.3	-	-	-	-	4.8	1627	0.4	8.3	-	-	-

Bias (%) = [(| Mean - Target |) / Target] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%)] / CV (%)

Month : 2025 12 Change

Cumulative : from 2021 01 01 to 2025 12 31 Change

TOP

Reagent Kit	Reagent Code
Medicon	5