

# QC Chart of Internal Quality Control (IQ)for G6PD Quantitative Test

Select LotNo :

## Lab019

QC Control Lot No.	Control N		Control D	
	BE0116N		BE0116D	
Duration of the Analyzing	Month (2026/01)	CUM (2025/08/20~2026/01/30)	Month (2026/01)	CUM (2025/08/20~2026/01/30)
Runs (N)	4	22	4	22
Mean (U/gHb)	13.5	13.5	4.8	4.8
SD	0.1	0.2	0.1	0.1
CV (%)	0.7	1.5	2.1	2.1
Target Value (U/gHb)	13.5	13.5	4.9	4.9
Total Error (%)	1.5	3.0	6.2	6.2
TEa (%)	20	20	20	20
$\sigma$	>6	>6	>6	>6

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%  
 TE : Total Error(%) = Bias (%) + 2 x CV (%)  
 $\sigma$  (Sigma) = [TEa% - Bias (%) ] / CV (%)


Month :    ; Cumulative : from    to

[\[TOP\]](#)

## Control N and Control D SDI QC Chart

Lot No.: BE0116N ; Duration : 2025-08-20 to 2026-01-30 ; Target : 13.5 ; SD : 0.56 ( ● )

Lot No.: BE0116D ; Duration : 2025-08-20 to 2026-01-30 ; Target : 4.9 ; SD : 0.32 ( □ )




Date 08-20 08-27 09-03 09-10 09-24 10-01 10-08 10-15 10-22 10-29 11-05 11-12 11-19 11-26 12-03 12-10 12-24 12-31 01-07 01-14 01-21 01-28

Month :    ; Cumulative : from    to

[\[TOP\]](#)

## Control N SDI QC Chart

Lot No.: BE0116N ; Duration : 2025-08-20 to 2026-01-30 ; Target : 13.5 ; SD : 0.56



Date 08-20 08-27 09-03 09-10 09-24 10-01 10-08 10-15 10-22 10-29 11-05 11-12 11-19 11-26 12-03 12-10 12-24 12-31 01-07 01-14 01-21 01-28

Month :    ; Cumulative : from    to

[\[TOP\]](#)

# Control D SDI QC Chart

Lab019

Lot No.: BE0116D ; Duration : 2025-08-20 to 2026-01-30 ; Target : 4.9 ; SD : 0.32



Date 08-20 08-27 09-03 09-10 09-24 10-01 10-08 10-15 10-22 10-29 11-05 11-12 11-19 11-26 12-03 12-10 12-24 12-31 01-07 01-14 01-21 01-28

Month : 2026 ▾ 01 ▾ Change ; Cumulative : from 2025 ▾ 08 ▾ 20 ▾ to 2026 ▾ 01 ▾ 30 ▾ Change

[TOP](#)

# Peer Group Statistics (Table 1)

Select LotNo : BE0116N (2025-02-12 ~ 2100-12-31) Change

Select Reagent Kit : 5 - Medicon Change

[Print Table 1](#)

## Monthly

Month : 2026 01 Change

UnitID	Reagent Kit (Code)	Control N (Lot No.: BE0116N)								Control D (Lot No.: BE0116D)							
		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab019</a>	5	13.5	13.5	4	0.1	0.7	1.5	20	>6	4.9	4.8	4	0.1	2.1	6.2	20	>6
<a href="#">Lab037</a>	5	13.5	13.0	2	-	-	-	20	-	4.9	4.7	2	-	-	-	20	-
<a href="#">Lab052</a>	5	13.5	13.3	2	-	-	-	20	-	4.9	4.8	2	-	-	-	20	-
Total	-	-	13.3	8	0.3	2.3	-	-	-	-	4.7	8	0.2	4.3	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[\[TOP\]](#)

## Cumulative

Cumulative : from 2025 04 01 to 2026 01 31 Change

UnitID	Reagent Kit (Code)	Control N (Lot No.: BE0116N)								Control D (Lot No.: BE0116D)							
		Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab018</a>	5	15.3	14.3	16	0.2	1.4	9.3	20	>6	4.7	5.1	16	0.1	2.0	12.4	20	5.7
<a href="#">Lab019</a>	5	13.5	13.5	22	0.2	1.5	3.0	20	>6	4.9	4.8	22	0.1	2.1	6.2	20	>6
<a href="#">Lab037</a>	5	13.5	13.0	2	-	-	-	20	-	4.9	4.7	2	-	-	-	20	-
<a href="#">Lab052</a>	5	13.5	13.4	6	0.2	1.5	3.7	20	>6	4.9	4.7	6	0.2	4.3	12.6	20	3.7
Total	-	-	13.8	46	0.5	3.6	-	-	-	-	4.9	46	0.2	4.1	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

[\[TOP\]](#)

Reagent Kit	Reagent Code
Medicon	5

## Peer Group Statistics (Table 2)

Select LotNo : BE0116N (2025-02-12 ~ 2100-12-31) v Change

Select Reagent Kit : 5 - Medicon v Change

[Print Table 2](#)

### Control N Month vs. Cumulative

		Control N (Lot No.: BE0116N)															
		Month (2026/01)								CUM (2025/04/01~2026/01/31)							
UnitID	Reagent Kit (Code)	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab018</a>	5	15.3	-	0	-	-	-	20	-	15.3	14.3	16	0.2	1.4	9.3	20	>6
<a href="#">Lab019</a>	5	13.5	13.5	4	0.1	0.7	1.5	20	>6	13.5	13.5	22	0.2	1.5	3.0	20	>6
<a href="#">Lab037</a>	5	13.5	13.0	2	-	-	-	20	-	13.5	13.0	2	-	-	-	20	-
<a href="#">Lab052</a>	5	13.5	13.3	2	-	-	-	20	-	13.5	13.4	6	0.2	1.5	3.7	20	>6
<b>Total</b>	-	-	13.3	8	0.3	2.3	-	-	-	-	13.8	46	0.5	3.6	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month : 2026 v 01 v Change    Cumulative : from 2025 v 04 v 01 v to 2026 v 01 v 31 v Change

[TOP](#)

### Control D Month vs. Cumulative

		Control D (Lot No.: BE0116D)															
		Month (2026/01)								CUM (2025/04/01~2026/01/31)							
UnitID	Reagent Kit (Code)	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ	Target (U/gHb)	Mean (U/gHb)	n for Mean	SD	CV (%)	TE (%)	TEa (%)	σ
<a href="#">Lab018</a>	5	4.7	-	0	-	-	-	20	-	4.7	5.1	16	0.1	2.0	12.4	20	5.7
<a href="#">Lab019</a>	5	4.9	4.8	4	0.1	2.1	6.2	20	>6	4.9	4.8	22	0.1	2.1	6.2	20	>6
<a href="#">Lab037</a>	5	4.9	4.7	2	-	-	-	20	-	4.9	4.7	2	-	-	-	20	-
<a href="#">Lab052</a>	5	4.9	4.8	2	-	-	-	20	-	4.9	4.7	6	0.2	4.3	12.6	20	3.7
<b>Total</b>	-	-	4.7	8	0.2	4.3	-	-	-	-	4.9	46	0.2	4.1	-	-	-

Bias (%) = [ ( | Mean - Target | ) / Target ] x 100%

TE : Total Error(%) = Bias (%) + 2 × CV (%)

σ (Sigma) = [TEa% - Bias (%) ] / CV (%)

Month : 2026 v 01 v Change    Cumulative : from 2025 v 04 v 01 v to 2026 v 01 v 31 v Change

[TOP](#)

Reagent Kit	Reagent Code
Medicon	5