

Summary Report of IQC program for G6PD Quantitative Test - AMP Group - November 2015 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	11	11
Received results number (n)	48	48
Median	12.6 (U/gHb)	3.0 (U/gHb)
Mean	12.7 (U/gHb)	3.1 (U/gHb)
SD	1.1	0.3
CV	8.7%	9.7%
Range of G6PD	10.6 ~ 14.9 (U/gHb)	2.5 ~ 3.9 (U/gHb)
Range of Hb	1.9 ~ 3.1 (g/dL)	2.2 ~ 3.7 (g/dL)

*The statistic results are calculated from all labs reported in this month

II. The distribution of G6PD reported for each lab in this survey

