

Summary Report of IQC program for G6PD Quantitative Test - AMP Group - December 2015 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	9	9
Received results number (n)	44	44
Median	12.8 (U/gHb)	3.1 (U/gHb)
Mean	12.9 (U/gHb)	3.2 (U/gHb)
SD	1.3	0.4
CV	10.1%	12.5%
Range of G6PD	9.8 ~ 14.9 (U/gHb)	2.2 ~ 3.8 (U/gHb)
Range of Hb	1.1 ~ 3.1 (g/dL)	1.4 ~ 3.4 (g/dL)

*The statistic results are calculated from all labs reported in this month

II. The distribution of G6PD reported for each lab in this survey

