

Summary Report of IQC program for G6PD Quantitative Test - AMP Group - January 2016 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	11	11
Received results number (n)	48	48
Median	12.4 (U/gHb)	3.0 (U/gHb)
Mean	12.6 (U/gHb)	3.0 (U/gHb)
SD	1.3	0.4
CV	10.3%	13.3%
Range of G6PD	9.8 ~ 14.8 (U/gHb)	2.2 ~ 3.9 (U/gHb)
Range of Hb	2.1 ~ 2.6 (g/dL)	2.2 ~ 3.1 (g/dL)

*The statistic results are calculated from all labs reported in this month

II. The distribution of G6PD reported for each lab in this survey

