

Summary Report of IQC program for G6PD Quantitative Test - AMP Group - August 2016 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	7	7
Received results number (n)	34	34
Median	12.8 (U/gHb)	3.0 (U/gHb)
Mean	12.6 (U/gHb)	3.0 (U/gHb)
SD	1.2	0.3
CV	9.5%	10.0%
Range of G6PD	9.9 ~ 14.4 (U/gHb)	2.4 ~ 3.5 (U/gHb)
Range of Hb	2.3 ~ 2.6 (g/dL)	2.7 ~ 3.1 (g/dL)

*The statistic results are calculated from all labs reported in this month

II. The distribution of G6PD reported for each lab in this survey

