

# Summary Report of IQC program for G6PD Quantitative Test - AMP Group - January 2017 -

## I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	8	8
Received results number (n)	29	29
Median	11.6 (U/gHb)	2.7 (U/gHb)
Mean	11.4 (U/gHb)	2.7 (U/gHb)
SD	1.6	0.4
CV	14.0%	14.8%
Range of G6PD	8.3 ~ 14.4 (U/gHb)	2.0 ~ 3.5 (U/gHb)
Range of Hb	2.3 ~ 3.6 (g/dL)	2.8 ~ 4.3 (g/dL)

\*The statistic results are calculated from all labs reported in this month

\*\* G6PD Method = AMP reagent kit, 37°C

## II. The distribution of G6PD reported for each lab in this survey

