

# Summary Report of IQC program for G6PD Quantitative Test - AMP Group - February 2017 -

## I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	6	6
Received results number (n)	26	26
Median	12.1 (U/gHb)	2.9 (U/gHb)
Mean	12.2 (U/gHb)	2.9 (U/gHb)
SD	1.2	0.3
CV	9.8%	10.3%
Range of G6PD	10.4 ~ 14.2 (U/gHb)	2.5 ~ 3.4 (U/gHb)
Range of Hb	2.5 ~ 3.4 (g/dL)	2.5 ~ 3.5 (g/dL)

\*The statistic results are calculated from all labs reported in this month

\*\* G6PD Method = AMP reagent kit, 37°C

## II. The distribution of G6PD reported for each lab in this survey

