

# Summary Report of IQC program for G6PD Quantitative Test - AMP Group - April 2017 -

## I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	10	10
Received results number (n)	46	46
Median	11.5 (U/gHb)	2.8 (U/gHb)
Mean	10.9 (U/gHb)	2.7 (U/gHb)
SD	1.9	0.4
CV	17.4%	14.8%
Range of G6PD	7.0 ~ 13.5 (U/gHb)	1.5 ~ 3.4 (U/gHb)
Range of Hb	2.0 ~ 3.8 (g/dL)	1.9 ~ 4.5 (g/dL)

\*The statistic results are calculated from all labs reported in this month

\*\* G6PD Method = AMP reagent kit, 37°C

## II. The distribution of G6PD reported for each lab in this survey

