
Summary Report of IQC program for G6PD Quantitative Test - Medicon Group - January 2021 -


I. The statistic results of all laboratories in this month


G6PD	Control N (Lot No.:AH1220N)	Control D (Lot No.:AH1220D)
Labs	9	9
Received results number (n)	50	50
Median	14.4 (U/gHb)	4.4 (U/gHb)
Mean	13.8 (U/gHb)	4.4 (U/gHb)
SD	1.4	0.3
CV	10.1%	6.8%
Range of G6PD	$10.5 \sim 16.0 (\text{U/gHb})$	$3.3 \sim 5.2 (\text{U/gHb})$
Range of Hb	$2.0 \sim 2.5 \text{ (g/dL)}$	$2.0 \sim 2.8 \; (g/dL)$

^{*}The statistic results are calculated from all labs reported in this month

II. The distribution of G6PD reported for each lab in this survey









^{**} G6PD Method = Medicon reagent kit, 37°C